


ISM02.4

LB 95909
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47830



A= 0.0000e+000

B= 2.1077e-004

C= -6.7897e-002

Rho= 0.9999137

Accept=Accepted

Slope
x-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.026	-0.026	200	0.000	200					-
0.2	0.200	0.175	-0.025	1153	0.0 %	1153					-13
2.5	2.500	2.554	0.054	12440	0.0 %	12440					2
5.0	5.000	5.061	0.061	24332	0.0 %	24332					1
7.5	7.500	7.422	-0.078	35537	0.0 %	35537					-1
10.0	10.000	10.013	0.013	47830	0.0 %	47830					0

LB
618118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	200	-		0 ISM02.4	PPB	6/8/2018 15:03	Std
	0.2 S0.2	1153	-		0.2 ISM02.4	PPB	6/8/2018 15:06	Std
	2.5 S2.5	12440	-		2.5 ISM02.4	PPB	6/8/2018 15:08	Std
	5 S5	24332	-		5 ISM02.4	PPB	6/8/2018 15:10	Std
	7.5 S7.5	35537	-		7.5 ISM02.4	PPB	6/8/2018 15:13	Std
	10 S10.0	47830	-		10 ISM02.4	PPB	6/8/2018 15:15	Std
ICV	ICV	19260	3.9916 -		ISM02.4	PPB	6/8/2018 15:18	SMPL
ICB	ICB	51	-0.0571 -		ISM02.4	PPB	6/8/2018 15:20	SMPL
CCV089	CCV089	24191	5.0309 -		ISM02.4	PPB	6/8/2018 15:22	SMPL
CCB089	CCB089	57	-0.0559 -		ISM02.4	PPB	6/8/2018 15:24	SMPL
PB110067BL	PBS004	116	-0.0434 -		ISM02.4	PPB	6/8/2018 15:27	SMPL
J3337-01	MET2Z3	7071	1.4225 -		ISM02.4	PPB	6/8/2018 15:29	SMPL
J3337-02	MET2Z4	8885	1.8048 -		ISM02.4	PPB	6/8/2018 15:31	SMPL
J3337-03	MET2Z5	6271	1.2539 -		ISM02.4	PPB	6/8/2018 15:34	SMPL
J3337-04	MET2Z6	11830	2.4256 -		ISM02.4	PPB	6/8/2018 15:36	SMPL
J3337-05	MET2Z7	3430	0.6551 -		ISM02.4	PPB	6/8/2018 15:38	SMPL
J3337-06	MET2Z8	16771	3.467 -		ISM02.4	PPB	6/8/2018 15:40	SMPL
J3337-07	MET2Z9	11927	2.446 -		ISM02.4	PPB	6/8/2018 15:43	SMPL
J3337-08	MET300	7566	1.5268 -		ISM02.4	PPB	6/8/2018 15:45	SMPL
J3337-09	MET300D	6482	1.2983 -		ISM02.4	PPB	6/8/2018 15:47	SMPL
J3337-10	MET300S	20082	4.1649 -		ISM02.4	PPB	6/8/2018 15:50	SMPL
J3337-11	MET301	3200	0.6066 -		ISM02.4	PPB	6/8/2018 15:52	SMPL
J3337-12	MET302	8783	1.7833 -		ISM02.4	PPB	6/8/2018 15:54	SMPL
J3337-13	MET303	20943	4.3463 -		ISM02.4	PPB	6/8/2018 15:56	SMPL
J3337-14	MET304	15771	3.2562 -		ISM02.4	PPB	6/8/2018 15:59	SMPL
J3337-15	MET305	2778	0.5176 -		ISM02.4	PPB	6/8/2018 16:01	SMPL
J3337-16	MET306	25190	5.2415 -		ISM02.4	PPB	6/8/2018 16:03	SMPL
J3337-17	MET307	7552	1.5239 -		ISM02.4	PPB	6/8/2018 16:05	SMPL
J3337-18	MET308	9286	1.8894 -		ISM02.4	PPB	6/8/2018 16:08	SMPL
CCV090	CCV090	24239	5.0411 -		ISM02.4	PPB	6/8/2018 16:10	SMPL
CCB090	CCB090	53	-0.0567 -		ISM02.4	PPB	6/8/2018 16:12	SMPL