



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95910

Slope : 98.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/08/2018	13:30
2	CAL2	1	Water			21.5	7.00	06/08/2018	13:32
3	CAL3	1	Water			21.7	10.03	06/08/2018	13:34
4	ICV	1	Water			21.5	7.00	06/08/2018	13:36
5	CCV1	1	Water			21.4	2.00	06/08/2018	13:38
6	J3390-01	1	Water			22.7	4.19	06/08/2018	13:40
7	J3390-01DUP	1	Water			22.8	4.20	06/08/2018	13:42
8	CCV2	1	Water			21.7	12.00	06/08/2018	13:44

WORKLIST(Hardcopy Internal Chain)

LB 95910

WorkList Name : pH060818

WorkList ID : 113133

Date : 6/8/2018 1:19:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2018	Water	J3390-01	pH	Cool 4 deg C	STER04	P11	POOL-WATER	06/07/2018	SM 4500-PH B

Date/Time 06/08/18 1:25 PM
 Received by: SP
 Relinquished by: CP

Date/Time 06/08/18 1:55 PM
 Received by: CP
 Relinquished by: SP