

LB95916


245.1

INSTRUMENT ID: LV2

Linear

 μ Abs.:

33991



A= 0.0000e+000

B= 2.9689e-004

C= -5.0341e-003

Rho= 0.9998803

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.012	0.012	58	0.000	58				
0.05	0.050	0.061	0.011	222	0.0 %	222				
0.2	0.200	0.194	-0.006	671	0.0 %	671				
2.5	2.500	2.527	0.027	8529	0.0 %	8529				
5.0	5.000	4.989	-0.011	16823	0.0 %	16823				
7.5	7.500	7.380	-0.120	24874	0.0 %	24874				
10.0	10.000	10.086	0.086	33991	0.0 %	33991				

LB95916 IONSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 08 Jun 2018 17:07:14

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	58	-	0.0000	245.1	PPB	08 Jun 2018 17:09:23	S	Std
0.05 - 1	50.05	222	-	0.0500	245.1	PPB	08 Jun 2018 17:11:31	S	Std
0.2 - 1	50.2	671	-	0.2000	245.1	PPB	08 Jun 2018 17:13:39	S	Std
2.5 - 1	52.5	8529	-	2.5000	245.1	PPB	08 Jun 2018 17:15:48	S	Std
5.0 - 1	55	16823	-	5.0000	245.1	PPB	08 Jun 2018 17:17:56	S	Std
7.5 - 1	57.5	24874	-	7.5000	245.1	PPB	08 Jun 2018 17:20:03	S	Std
10.0 - 1	51.6	33991	-	10.0000	245.1	PPB	08 Jun 2018 17:26:02	S	Std
ICV92 - 1	ICV92	12933	3.8346	-245.1		PPB	08 Jun 2018 17:28:53	U	SMPL
ICB92 - 1	ICB92	-67	-0.0249	-245.1		PPB	08 Jun 2018 17:30:59	U	SMPL
CCV91 - 1	CCV91	17008	5.0444	-245.1		PPB	08 Jun 2018 17:33:06	U	SMPL
CCB91 - 1	CCB91	-79	-0.0285	-245.1		PPB	08 Jun 2018 17:35:12	U	SMPL
CRA - 1	CRA	191	0.0517	-245.1		PPB	08 Jun 2018 17:37:20	U	SMPL
HighStd - 1	HighStd	33444	9.9240	-245.1		PPB	08 Jun 2018 17:39:26	U	SMPL
ChkStd - 1	ChkStd	23496	6.9706	-245.1		PPB	08 Jun 2018 17:41:33	U	SMPL
PB110060BL - 1	PBW	-50	-0.0199	-245.1		PPB	08 Jun 2018 17:43:43	U	SMPL
PB110060BS - 1	LCSW	13951	4.1368	-245.1		PPB	08 Jun 2018 17:45:51	U	SMPL
J3334-01 - 1	COMP	75	0.0172	-245.1		PPB	08 Jun 2018 17:47:59	U	SMPL
J3343-01 - 1	060518-DBRC-E-D-B	183	0.0493	-245.1		PPB	08 Jun 2018 17:50:07	U	SMPL
J3343-02 - 1	060518-DBRC-E-D-S	187	0.0505	-245.1		PPB	08 Jun 2018 17:52:15	U	SMPL
J3343-03 - 1	060518-DBRC-E-U-B	206	0.0561	-245.1		PPB	08 Jun 2018 17:54:23	U	SMPL
J3343-04 - 1	060518-DBRC-E-U-S	201	0.0546	-245.1		PPB	08 Jun 2018 17:56:30	U	SMPL
CCV92 - 1	CCV92	16855	4.9990	-245.1		PPB	08 Jun 2018 17:58:37	U	SMPL
CCB92 - 1	CCB92	-80	-0.0288	-245.1		PPB	08 Jun 2018 18:00:44	U	SMPL
J3343-04DUP - 1	060518-DBRC-E-U-SDUP	162	0.0431	-245.1		PPB	08 Jun 2018 18:02:52	U	SMPL
J3343-04MS - 1	060518-DBRC-E-U-SMS	15912	4.7190	-245.1		PPB	08 Jun 2018 18:04:58	U	SMPL
J3343-04MSD - 1	060518-DBRC-E-U-SMSD	16204	4.8057	-245.1		PPB	08 Jun 2018 18:07:05	U	SMPL
J3343-04LX5 - 1		-25	-0.0125	-245.1		PPB	08 Jun 2018 18:09:12	U	SMPL
J3343-04A - 1		16687	4.9491	-245.1		PPB	08 Jun 2018 18:11:20	U	SMPL
CCV93 - 1	CCV93	17054	5.0581	-245.1		PPB	08 Jun 2018 18:13:27	U	SMPL
CCB93 - 1	CCB93	-103	-0.0356	-245.1		PPB	08 Jun 2018 18:15:36	U	SMPL