

LB95923


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

35614



A= 0.0000e+000

B= 2.7650e-004

C= 1.3410e-002

Rho= 0.9985859

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.021	0.021	26	0.000	26				
0.2	0.200	0.203	0.003	687	0.0 %	687				
2.5	2.500	2.594	0.094	9334	0.0 %	9334				
5.0	5.000	4.693	-0.307	16924	0.0 %	16924				
7.5	7.500	7.828	0.328	28264	0.0 %	28264				
10.0	10.000	9.861	-0.139	35614	0.0 %	35614				

LB95923 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 11 Jun 2018 10:01:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	26	-	0.0000	7470A	PPB	11 Jun 2018 10:07:04	S	Std
0.2 - 1	50.2	687	-	0.2000	7470A	PPB	11 Jun 2018 10:09:12	S	Std
2.5 - 1	52.5	9334	-	2.5000	7470A	PPB	11 Jun 2018 10:11:20	S	Std
5.0 - 1	55	16924	-	5.0000	7470A	PPB	11 Jun 2018 10:13:29	S	Std
7.5 - 1	57.5	28264	-	7.5000	7470A	PPB	11 Jun 2018 10:15:36	S	Std
10.0 - 1	510	35614	-	10.0000	7470A	PPB	11 Jun 2018 10:20:14	S	Std
ICV93 - 1	ICV93	14224	3.9463	-	7470A	PPB	11 Jun 2018 10:23:03	U	SMPL
ICB93 - 1	ICB93	-115	-0.0184	-	7470A	PPB	11 Jun 2018 10:25:09	U	SMPL
CCV94 - 1	CCV94	17023	4.7202	-	7470A	PPB	11 Jun 2018 10:27:15	U	SMPL
CCB94 - 1	CCB94	-122	-0.0203	-	7470A	PPB	11 Jun 2018 10:29:22	U	SMPL
CRA - 1	CRA	672	0.1992	-	7470A	PPB	11 Jun 2018 10:31:28	U	SMPL
HighStd - 1	HighStd	36108	9.9972	-	7470A	PPB	11 Jun 2018 10:33:35	U	SMPL
ChkStd - 1	ChkStd	22900	6.3452	-	7470A	PPB	11 Jun 2018 10:38:46	U	SMPL
PB110094BL - 1	PBW	-164	-0.0319	-	7470A	PPB	11 Jun 2018 10:40:53	U	SMPL
PB110094BS - 1	LCSW	13635	3.7835	-	7470A	PPB	11 Jun 2018 10:43:02	U	SMPL
PB110094TB - 1	PB110062TB	-68	-0.0054	-	7470A	PPB	11 Jun 2018 10:45:09	U	SMPL
J3240-02 - 1	BU-02-060718	-8	0.0112	-	7470A	PPB	11 Jun 2018 10:47:17	U	SMPL
J3240-02DUP - 1	BU-02-060718DUP	10	0.0162	-	7470A	PPB	11 Jun 2018 10:49:25	U	SMPL
J3240-02MS - 1	BU-02-060718MS	13667	3.7923	-	7470A	PPB	11 Jun 2018 10:51:34	U	SMPL
J3240-02MSD - 1	BU-02-060718MSD	13254	3.6781	-	7470A	PPB	11 Jun 2018 10:53:41	U	SMPL
CCV95 - 1	CCV95	17313	4.8004	-	7470A	PPB	11 Jun 2018 10:55:47	U	SMPL
CCB95 - 1	CCB95	-81	-0.0090	-	7470A	PPB	11 Jun 2018 10:57:54	U	SMPL
J3247-02 - 1	OR-02-060718	19	0.0187	-	7470A	PPB	11 Jun 2018 11:00:01	U	SMPL
J3360-03 - 1	JC-SOIL-COMP	13	0.0170	-	7470A	PPB	11 Jun 2018 11:02:07	U	SMPL
J3368-05 - 1	S-1	10	0.0162	-	7470A	PPB	11 Jun 2018 11:04:14	U	SMPL
J3377-02 - 1	TP-B-DB-S	44	0.0256	-	7470A	PPB	11 Jun 2018 11:06:21	U	SMPL
J3377-05 - 1	TP-B-D	33	0.0225	-	7470A	PPB	11 Jun 2018 11:08:28	U	SMPL
J3377-09 - 1	TP-C-DB	48	0.0267	-	7470A	PPB	11 Jun 2018 11:10:36	U	SMPL
J3240-02LX5 - 1		7	0.0153	-	7470A	PPB	11 Jun 2018 11:12:44	U	SMPL
J3240-02A - 1		14810	4.1083	-	7470A	PPB	11 Jun 2018 11:14:52	U	SMPL
CCV96 - 1	CCV96	16964	4.7039	-	7470A	PPB	11 Jun 2018 11:17:01	U	SMPL
CCB96 - 1	CCB96	-95	-0.0129	-	7470A	PPB	11 Jun 2018 11:19:08	U	SMPL