



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95928

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	06/11/2018	09:10
2	J3414-01		1	>100.0	>212.0	>212.0	06/11/2018	09:40
3	J3414-01DUP		1	>100.0	>212.0	>212.0	06/11/2018	10:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB95928

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-061118

WorkList ID : 113164

Date : 6/11/2018 8:02:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2018	Solid	J3414-01	B Flash Point	Cool 4 deg C	GROU11	H22	PGWPAL-OIL-060618	06/06/2018	1010A

Date/Time 06/11/18 8:30 AM
Received by: KS
Relinquished by: CP

Date/Time 06/11/18 10:15 AM
Received by: CP
Relinquished by: KS