



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95931

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	06/11/2018	13:45
2	CAL2	1	Water			21.6	7.00	06/11/2018	13:47
3	CAL3	1	Water			21.5	10.03	06/11/2018	13:49
4	ICV	1	Water			21.6	7.00	06/11/2018	13:51
5	CCV1	1	Water			21.4	2.00	06/11/2018	13:53
6	J3425-01	1	Water			22.3	5.77	06/11/2018	13:55
7	J3425-01DUP	1	Water			22.4	5.78	06/11/2018	13:57
8	CCV2	1	Water			21.7	12.00	06/11/2018	13:59

WORKLIST(Hardcopy Internal Chain)

LB95931

WorkList Name : corrosivity j3425

WorkList ID : 113189

Date : 6/11/2018 1:36:11 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2018	Water	J3425-01	Corrosivity	Cool 4 deg C	GROU11	B52	PGWPAL-AQ060818	06/08/2018	9040C

Date/Time 06/11/18 1:40 PM
 Received by: SP
 Relinquished by: CP

Date/Time 06/11/18 2:05 PM
 Received by: CP
 Relinquished by: SP