



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95936

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	06/11/2018	17:10
2	CAL2	1	Water			21.5	7.00	06/11/2018	17:12
3	CAL3	1	Water			21.6	10.03	06/11/2018	17:14
4	ICV	1	Water			21.5	7.00	06/11/2018	17:16
5	CCV1	1	Water			21.6	2.00	06/11/2018	17:18
6	J2979-03	1	Solid	20.02	20	23.5	9.03	06/11/2018	17:20
7	J2979-03DUP	1	Solid	20.04	20	23.6	9.04	06/11/2018	17:22
8	J3245-02	1	Solid	20.02	20	24.4	8.37	06/11/2018	17:24
9	J3245-04	1	Solid	20.03	20	24.1	7.91	06/11/2018	17:26
10	J3245-06	1	Solid	20.02	20	21.4	8.63	06/11/2018	17:28
11	J3245-08	1	Solid	20.04	20	23.9	7.94	06/11/2018	17:30
12	J3247-04	1	Solid	20.02	20	24.2	7.87	06/11/2018	17:32
13	J3429-03	1	Solid	20.03	20	23.7	5.73	06/11/2018	17:34
14	J3442-02	1	Solid	20.03	20	23.1	8.05	06/11/2018	17:36
15	CCV2	1	Water			21.6	12.00	06/11/2018	17:38

WORKLIST(Hardcopy Internal Chain)

LB95936

WorkList Name : corrosivity061118

WorkList ID : 113199

Date : 6/11/2018 3:56:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/11/2018	Solid	J2979-03	Corrosivity	Cool 4 deg C	PSEG01	I63	STONE-DUST-CANVAS-BAG	06/11/2018	9045C
06/11/2018	Solid	J3245-02	Corrosivity	Cool 4 deg C	PSEG05	I63	CL-01-061118-A	06/11/2018	9045C
06/11/2018	Solid	J3245-04	Corrosivity	Cool 4 deg C	PSEG05	I63	CL-01-061118-B	06/11/2018	9045C
06/11/2018	Solid	J3245-06	Corrosivity	Cool 4 deg C	PSEG05	I63	CL-01-061118-C	06/11/2018	9045C
06/11/2018	Solid	J3245-08	Corrosivity	Cool 4 deg C	PSEG05	I63	CL-01-061118-D	06/11/2018	9045C
06/11/2018	Solid	J3247-04	Corrosivity	Cool 4 deg C	PSEG05		OR-03-061118	06/11/2018	9045C
06/11/2018	Solid	J3429-03	Corrosivity	Cool 4 deg C	PSEG01	I63	OILY-STONES	06/11/2018	9045C
06/11/2018	Solid	J3442-02	Corrosivity	Cool 4 deg C	PARS15	P51	CAS1-WC-5-01-1-57	06/11/2018	9045C

Date/Time 06/11/18 05:00 PM

Received by: SP

Relinquished by: CP

Date/Time 06/11/18 05:50 PM

Received by: CP

Relinquished by: SP