



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: APARNADate: 06/12/2018TEMP1 IN: 103 06/12/2018 07:20 TEMP OUT: 104 06/12/2018 08:30TEMP2 IN: 104 06/12/2018 09:20 TEMP OUT: 104 06/12/2018 10:25TEMP3 IN: 105 06/12/2018 12:30 TEMP OUT: 105 06/12/2018 14:00TEMP4 IN: 105 06/12/2018 14:40 TEMP OUT: 105 06/12/2018 16:25Run Number: LB95941BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95941BL	LB95941BL	1.3905	1.3905	100	1.3905	1.3905	1.3905	0.0000	0
2	LB95941BS	LB95941BS	1.4031	1.4031	100	1.4530	1.4530	1.4530	0.0499	499
3	J3433-01	060818-JETT-E-D-B	1.3850	1.3850	200	1.3956	1.3956	1.3956	0.0106	53
4	J3433-01DUP	060818-JETT-E-D-BDUP	1.4036	1.4036	200	1.4143	1.4143	1.4143	0.0107	53.5
5	J3433-02	060818-JETT-E-D-M	1.3942	1.3942	400	1.3997	1.3997	1.3997	0.0055	13.7
6	J3433-03	060818-JETT-E-D-S	1.4123	1.4123	400	1.4192	1.4192	1.4192	0.0069	17.3
7	J3433-04	060818-JETT-E-U-B	1.4106	1.4106	400	1.4323	1.4323	1.4323	0.0217	54.3
8	J3433-05	060818-JETT-E-U-M	1.4045	1.4045	400	1.4153	1.4153	1.4153	0.0108	27
9	J3433-06	060818-JETT-E-U-S	1.3856	1.3856	400	1.3911	1.3911	1.3911	0.0055	13.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061218-2

WorkList ID : 113213

Date : 6/12/2018 7:59:09 AM

2095941

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2018	Water	J3433-01 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-E-D-B	06/08/2018	SM2540 D
06/15/2018	Water	J3433-02 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-E-D-M	06/08/2018	SM2540 D
06/15/2018	Water	J3433-03 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-E-D-S	06/08/2018	SM2540 D
06/15/2018	Water	J3433-04 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-E-U-B	06/08/2018	SM2540 D
06/15/2018	Water	J3433-05 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-E-U-M	06/08/2018	SM2540 D
06/15/2018	Water	J3433-06 C	TSS	Cool 4 deg C	HDR103	P13	060818-JETT-U-U-S	06/08/2018	SM2540 D

Date/Time 06/12/18, 8:20
 Received by: AS
 Relinquished by: CS

Date/Time 06/12/18, 13:00
 Received by: AS
 Relinquished by: AS