



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: KETANKUMAR

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95947

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.6	06/12/2018	09:20
2	J3425-01		1	>100.0	>212.0	>212.0	06/12/2018	09:45
3	J3425-01DUP		1	>100.0	>212.0	>212.0	06/12/2018	10:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB95947

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-061218

WorkList ID : 113226

Date : 6/12/2018 8:36:12 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/22/2018	Water	J3425-01	Flash Point	Cool 4 deg C	GROU11	B52	PGWPAL-AQ060818	06/08/2018	1010A

Date/Time 06/12/18 08:45 AM

Received by: KS

Relinquished by: CP

Date/Time 06/12/18 10:20 AM

Received by: CP

Relinquished by: KS