


245.1

LB95951
INSTRUMENT ID: CV1

Linear

 μ Abs.:

33255



A= 0.0000e+000

B= 3.0079e-004

C= -3.0564e-002

Rho= 0.9999861

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	91	0.000	91				
0.05	0.050	0.053	0.003	278	0.0 %	278				
0.2	0.200	0.200	0.000	767	0.0 %	767				
2.5	2.500	2.486	-0.014	8368	0.0 %	8368				
5.0	5.000	5.000	0.000	16726	0.0 %	16726				
7.5	7.500	7.541	0.041	25173	0.0 %	25173				
10.0	10.000	9.972	-0.028	33255	0.0 %	33255				

LB95951 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Jun 2018 13:46:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	91	-	0.0000	245.1	PPB	12 Jun 2018 13:48:40	S	Std
0.05 - 1	50.05	278	-	0.0500	245.1	PPB	12 Jun 2018 13:50:48	S	Std
0.2 - 1	50.2	767	-	0.2000	245.1	PPB	12 Jun 2018 13:52:57	S	Std
2.5 - 1	52.5	8368	-	2.5000	245.1	PPB	12 Jun 2018 13:55:05	S	Std
5.0 - 1	55	16726	-	5.0000	245.1	PPB	12 Jun 2018 13:57:13	S	Std
7.5 - 1	57.5	25173	-	7.5000	245.1	PPB	12 Jun 2018 13:59:20	S	Std
10.0 - 1	510	33255	-	10.0000	245.1	PPB	12 Jun 2018 14:01:28	S	Std
ICV96 - 1	ICV96	13668	4.0806	-245.1		PPB	12 Jun 2018 14:06:32	U	SMPL
ICB96 - 1	ICB96	-23	-0.0375	-245.1		PPB	12 Jun 2018 14:08:38	U	SMPL
CCV05 - 1	CCV05	16799	5.0223	-245.1		PPB	12 Jun 2018 14:10:45	U	SMPL
CCB05 - 1	CCB05	-56	-0.0474	-245.1		PPB	12 Jun 2018 14:12:51	U	SMPL
CRA - 1	CRA	251	0.0449	-245.1		PPB	12 Jun 2018 14:14:59	U	SMPL
HighStd - 1	HighStd	33458	10.0331	-245.1		PPB	12 Jun 2018 14:17:05	U	SMPL
ChkStd - 1	ChkStd	23941	7.1706	-245.1		PPB	12 Jun 2018 14:19:12	U	SMPL
PB110135BL - 1	PBW	50	-0.0155	-245.1		PPB	12 Jun 2018 14:24:01	U	SMPL
PB110135BS - 1	LCSW	12656	3.7762	-245.1		PPB	12 Jun 2018 14:26:08	U	SMPL
J3334-01 - 1	COMP	-15	-0.0351	-245.1		PPB	12 Jun 2018 14:28:16	U	SMPL
J3343-01 - 1	060518-DBRC-E-D-B	7	-0.0285	-245.1		PPB	12 Jun 2018 14:30:24	U	SMPL
J3343-02 - 1	060518-DBRC-E-D-S	26	-0.0227	-245.1		PPB	12 Jun 2018 14:32:33	U	SMPL
J3343-03 - 1	060518-DBRC-E-U-B	39	-0.0188	-245.1		PPB	12 Jun 2018 14:34:41	U	SMPL
J3343-04 - 1	060518-DBRC-E-U-S	68	-0.0101	-245.1		PPB	12 Jun 2018 14:36:48	U	SMPL
CCV06 - 1	CCV06	16781	5.0169	-245.1		PPB	12 Jun 2018 14:38:55	U	SMPL
CCB06 - 1	CCB06	-48	-0.0450	-245.1		PPB	12 Jun 2018 14:41:02	U	SMPL
J3411-01 - 1	060618-JETT-E-D-B	108	0.0019	-245.1		PPB	12 Jun 2018 14:43:10	U	SMPL
J3411-02 - 1	060618-JETT-E-D-M	97	-0.0014	-245.1		PPB	12 Jun 2018 14:45:16	U	SMPL
J3411-03 - 1	060618-JETT-E-D-S	101	-0.0002	-245.1		PPB	12 Jun 2018 14:47:23	U	SMPL
J3411-04 - 1	060618-JETT-E-U-B	44	-0.0173	-245.1		PPB	12 Jun 2018 14:49:30	U	SMPL
J3411-05 - 1	060618-JETT-E-U-M	19	-0.0248	-245.1		PPB	12 Jun 2018 14:51:38	U	SMPL
J3411-06 - 1	060618-JETT-E-U-S	44	-0.0173	-245.1		PPB	12 Jun 2018 14:53:46	U	SMPL
J3411-07 - 1	060618-DBRI-F-D-B	49	-0.0158	-245.1		PPB	12 Jun 2018 14:55:54	U	SMPL
J3411-08 - 1	060618-DBRI-F-D-S	57	-0.0134	-245.1		PPB	12 Jun 2018 14:58:02	U	SMPL
J3411-09 - 1	060618-DBRI-F-U-B	76	-0.0077	-245.1		PPB	12 Jun 2018 15:00:11	U	SMPL
J3411-10 - 1	060618-DBRI-F-U-S	86	-0.0047	-245.1		PPB	12 Jun 2018 15:02:18	U	SMPL
CCV07 - 1	CCV07	16889	5.0494	-245.1		PPB	12 Jun 2018 15:04:25	U	SMPL
CCB07 - 1	CCB07	-30	-0.0396	-245.1		PPB	12 Jun 2018 15:06:32	U	SMPL
J3411-10DUP - 1	060618-DBRI-F-U-SDUP	66	-0.0107	-245.1		PPB	12 Jun 2018 15:08:40	U	SMPL
J3411-10MS - 1	060618-DBRI-F-U-SMS	15345	4.5850	-245.1		PPB	12 Jun 2018 15:10:46	U	SMPL
J3411-10MSD - 1	060618-DBRI-F-U-SMSD	15243	4.5543	-245.1		PPB	12 Jun 2018 15:12:54	U	SMPL
PB110136BL - 1	PBW	-103	-0.0615	-245.1		PPB	12 Jun 2018 15:15:01	U	SMPL
PB110136BS - 1	LCSW	13587	4.0562	-245.1		PPB	12 Jun 2018 15:17:08	U	SMPL
J3412-01 - 1	060718-DBRI-E-D-B	-79	-0.0543	-245.1		PPB	12 Jun 2018 15:19:16	U	SMPL
J3412-02 - 1	060718-DBRI-E-D-S	28	-0.0221	-245.1		PPB	12 Jun 2018 15:21:24	U	SMPL
J3412-03 - 1	060718-DBRI-E-U-B	16	-0.0258	-245.1		PPB	12 Jun 2018 15:23:32	U	SMPL
J3412-04 - 1	060718-DBRI-E-U-S	23	-0.0236	-245.1		PPB	12 Jun 2018 15:25:41	U	SMPL
J3412-05 - 1	060718-JETT-E-D-B	37	-0.0194	-245.1		PPB	12 Jun 2018 15:27:48	U	SMPL
CCV08 - 1	CCV08	16937	5.0638	-245.1		PPB	12 Jun 2018 15:29:55	U	SMPL
CCB08 - 1	CCB08	-49	-0.0453	-245.1		PPB	12 Jun 2018 15:32:03	U	SMPL
J3412-06 - 1	060718-JETT-E-D-M	4	-0.0294	-245.1		PPB	12 Jun 2018 15:34:10	U	SMPL
J3412-07 - 1	060718-JETT-E-D-S	22	-0.0239	-245.1		PPB	12 Jun 2018 15:36:18	U	SMPL
J3412-08 - 1	060718-JETT-E-U-B	26	-0.0227	-245.1		PPB	12 Jun 2018 15:38:25	U	SMPL
J3412-09 - 1	060718-JETT-E-U-M	31	-0.0212	-245.1		PPB	12 Jun 2018 15:40:33	U	SMPL
J3412-10 - 1	060718-JETT-E-U-S	19	-0.0248	-245.1		PPB	12 Jun 2018 15:42:40	U	SMPL
J3412-10DUP - 1	060718-JETT-E-U-SDUP	26	-0.0227	-245.1		PPB	12 Jun 2018 15:44:48	U	SMPL
J3412-10MS - 1	060718-JETT-E-U-SMS	15802	4.7225	-245.1		PPB	12 Jun 2018 15:46:55	U	SMPL
J3412-10MSD - 1	060718-JETT-E-U-SMSD	15163	4.5303	-245.1		PPB	12 Jun 2018 15:49:03	U	SMPL
J3433-01 - 1	060818-JETT-E-D-B	-65	-0.0501	-245.1		PPB	12 Jun 2018 15:51:10	U	SMPL
J3433-02 - 1	060818-JETT-E-D-M	32	-0.0209	-245.1		PPB	12 Jun 2018 15:53:18	U	SMPL
CCV09 - 1	CCV09	16639	4.9742	-245.1		PPB	12 Jun 2018 15:55:25	U	SMPL
CCB09 - 1	CCB09	-59	-0.0483	-245.1		PPB	12 Jun 2018 15:57:33	U	SMPL
J3433-03 - 1	060818-JETT-E-D-S	11	-0.0273	-245.1		PPB	12 Jun 2018 15:59:40	U	SMPL
J3433-04 - 1	060818-JETT-E-U-B	88	-0.0041	-245.1		PPB	12 Jun 2018 16:01:48	U	SMPL
J3433-05 - 1	060818-JETT-E-U-M	99	-0.0008	-245.1		PPB	12 Jun 2018 16:03:55	U	SMPL
J3433-06 - 1	060818-JETT-E-U-S	108	0.0019	-245.1		PPB	12 Jun 2018 16:06:03	U	SMPL
J3411-10LX5 - 1		68	-0.0101	-245.1		PPB	12 Jun 2018 16:08:10	U	SMPL
J3411-10A - 1		16472	4.9240	-245.1		PPB	12 Jun 2018 16:10:18	U	SMPL
J3412-10LX5 - 1		-42	-0.0432	-245.1		PPB	12 Jun 2018 16:12:25	U	SMPL
J3412-10A - 1		16553	4.9483	-245.1		PPB	12 Jun 2018 16:14:33	U	SMPL
CCV10 - 1	CCV10	16715	4.9971	-245.1		PPB	12 Jun 2018 16:16:40	U	SMPL
CCB10 - 1	CCB10	-50	-0.0456	-245.1		PPB	12 Jun 2018 16:18:48	U	SMPL