



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95961

Slope : 99.2

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/12/2018	11:20
2	CAL2	1	Water			21.7	7.00	06/12/2018	11:22
3	CAL3	1	Water			21.5	10.03	06/12/2018	11:24
4	ICV	1	Water			21.5	7.00	06/12/2018	11:26
5	CCV1	1	Water			21.6	2.00	06/12/2018	11:28
6	J3445-01	1	Water			19.9	4.11	06/12/2018	11:30
7	J3445-01DUP	1	Water			20.0	4.12	06/12/2018	11:32
8	J3450-01	1	Water			22.4	5.40	06/12/2018	11:34
9	CCV2	1	Water			21.6	12.00	06/12/2018	11:36

WORKLIST(Hardcopy Internal Chain)

LB95961

WorkList Name : pH J3445

WorkList ID : 113235

Date : 6/12/2018 10:59:22 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/07/2018	Water	J3445-01	pH	Cool 4 deg C	STER05	P22	MINI-CELL	06/07/2018	SM 4500-PH B
06/11/2018	Water	J3450-01	pH	Cool 4 deg C	STER01	P53	POOL-WATER	06/11/2018	SM 4500-PH B

Date/Time 06/12/18 11:10 AM
 Received by: SP
 Relinquished by: CP

Date/Time 06/12/18 11:50 AM
 Received by: CP
 Relinquished by: SP