



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95969

Slope : 98.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/12/2018	15:50
2	CAL2	1	Water			21.7	7.00	06/12/2018	15:52
3	CAL3	1	Water			21.6	10.03	06/12/2018	15:54
4	ICV	1	Water			21.5	7.00	06/12/2018	15:56
5	CCV1	1	Water			21.6	2.00	06/12/2018	15:58
6	J3455-01	1	Water			17.9	4.40	06/12/2018	16:00
7	J3455-01DUP	1	Water			18.0	4.41	06/12/2018	16:02
8	J3455-02	1	Water			18.4	4.70	06/12/2018	16:04
9	J3455-03	1	Water			19.0	4.85	06/12/2018	16:06
10	J3455-04	1	Water			19.5	4.97	06/12/2018	16:08
11	CCV2	1	Water			21.6	12.00	06/12/2018	16:10

LB95969

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH061218

WorkList ID : 113248

Date : 6/12/2018 3:13:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2018	Water	J3455-01	pH	Cool 4 deg C	NYUL01	D41	TH-430-DI1	06/12/2018	SM 4500-PH B
06/12/2018	Water	J3455-02	pH	Cool 4 deg C	NYUL01	D41	TH-430-DI2	06/12/2018	SM 4500-PH B
06/12/2018	Water	J3455-03	pH	Cool 4 deg C	NYUL01	D41	TH-430-DI3	06/12/2018	SM 4500-PH B
06/12/2018	Water	J3455-04	pH	Cool 4 deg C	NYUL01	D41	TH-404-DI4	06/12/2018	SM 4500-PH B

Date/Time 06/12/18 3:40 PM
Received by: SP
Relinquished by: CP

Date/Time 06/12/18 4:20 PM
Received by: CP
Relinquished by: SP