



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95971

Slope : 98.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/12/2018	15:50
2	CAL2	1	Water			21.7	7.00	06/12/2018	15:52
3	CAL3	1	Water			21.6	10.03	06/12/2018	15:54
4	ICV	1	Water			21.5	7.00	06/12/2018	15:56
5	CCV1	1	Water			21.6	2.00	06/12/2018	17:45
6	J3448-01	1	Water			23.2	7.51	06/12/2018	17:47
7	J3448-01DUP	1	Water			23.2	7.52	06/12/2018	17:49
8	J3448-03	1	Water			23.4	7.52	06/12/2018	17:51
9	J3448-05	1	Water			23.1	7.50	06/12/2018	17:53
10	J3451-01	1	Water			23.9	7.77	06/12/2018	17:55
11	J3451-03	1	Water			23.6	7.57	06/12/2018	17:57
12	J3456-01	1	Water			25.6	6.66	06/12/2018	17:59
13	CCV2	1	Water			21.9	12.00	06/12/2018	18:01

WORKLIST(Hardcopy Internal Chain)

LB 95 971

WorkList Name : pH061218-1

WorkList ID : 113269

Date : 6/12/2018 3:13:34 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2018	Water	J3448-01	pH	Cool 4 deg C	PSEG01	D11	N-48971	06/12/2018	9040C
06/12/2018	Water	J3448-03	pH	Cool 4 deg C	PSEG01	D11	N-48969	06/12/2018	9040C
06/12/2018	Water	J3448-05	pH	Cool 4 deg C	PSEG01	D11	N-48966	06/12/2018	9040C
06/12/2018	Water	J3451-01	pH	Cool 4 deg C	PSEG01	D11	VB-25830-SLUDGEBOX	06/12/2018	9040C
06/12/2018	Water	J3451-03	pH	Cool 4 deg C	PSEG01	D11	VB-44949-SLUDGEBOX	06/12/2018	9040C
06/12/2018	Water	J3456-01	pH	Cool 4 deg C	PSEG01	D12	ES-OILY-WATER	06/12/2018	9040C

Date/Time 06/12/18 5:30 PM

Received by: SP

Relinquished by: CP

Date/Time 06/12/18 6:10 PM

Received by: CP

Relinquished by: SP