



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95973

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/12/2018	15:50
2	CAL2	1	Water			21.7	7.00	06/12/2018	15:52
3	CAL3	1	Water			21.6	10.03	06/12/2018	15:54
4	ICV	1	Water			21.5	7.00	06/12/2018	15:56
5	CCV1	1	Water			21.6	2.00	06/12/2018	18:05
6	J3246-02	1	Solid	20.03	20	25.2	8.30	06/12/2018	18:07
7	J3246-02DUP	1	Solid	20.02	20	25.3	8.31	06/12/2018	18:09
8	J3251-04	1	Solid	20.04	20	25.4	7.12	06/12/2018	18:11
9	J3459-02	1	Solid	20.02	20	25.3	6.09	06/12/2018	18:13
10	CCV2	1	Water			21.8	12.00	06/12/2018	18:15

LB 95973

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity061218-1

WorkList ID : 113255

Date : 6/12/2018 4:49:28 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2018	Solid	J3246-02	Corrosivity	Cool 4 deg C	PSEG05	D11	OK-02-061218	06/12/2018	9045C
06/12/2018	Solid	J3251-04	Corrosivity	Cool 4 deg C	PSEG05		SU-03-061118	06/12/2018	9045C
06/12/2018	Solid	J3459-02	Corrosivity	Cool 4 deg C	PSEG01	D12	MH-RO-341	06/12/2018	9045C

Date/Time 06/12/18 5:30 PM

Received by: SP

Relinquished by: CP

Date/Time 06/12/18 6:25 PM

Received by: CP

Relinquished by: SP