


7470A

LB95977
INSTRUMENT ID: C44

Linear

 μ Abs.:

38294



A= 0.0000e+000

B= 2.6418e-004

C= -1.0645e-001

Rho= 0.9999065

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.015	-0.015	346	0.000	346				
0.2	0.200	0.151	-0.049	973	0.0 %	973				
2.5	2.500	2.589	0.089	10202	0.0 %	10202				
5.0	5.000	5.028	0.028	19436	0.0 %	19436				
7.5	7.500	7.438	-0.062	28558	0.0 %	28558				
10.0	10.000	10.010	0.010	38294	0.0 %	38294				

LB95977 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Jun 2018 10:35:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	346	-	0.0000	7470A	PPB	13 Jun 2018 10:39:07	S	Std
0.2 - 1	50.2	973	-	0.2000	7470A	PPB	13 Jun 2018 10:41:16	S	Std
2.5 - 1	52.5	10202	-	2.5000	7470A	PPB	13 Jun 2018 10:43:24	S	Std
5.0 - 1	55	19436	-	5.0000	7470A	PPB	13 Jun 2018 10:45:33	S	Std
7.5 - 1	57.5	28558	-	7.5000	7470A	PPB	13 Jun 2018 10:47:41	S	Std
10.0 - 1	510	38294	-	10.0000	7470A	PPB	13 Jun 2018 10:53:06	S	Std
ICV98 - 1	ICV98	14196	3.6438	-	7470A	PPB	13 Jun 2018 10:55:41	U	SMPL
ICB98 - 1	ICB98	208	-0.0515	-	7470A	PPB	13 Jun 2018 10:57:47	U	SMPL
CCV16 - 1	CCV16	19624	5.0777	-	7470A	PPB	13 Jun 2018 10:59:54	U	SMPL
CCB16 - 1	CCB16	211	-0.0507	-	7470A	PPB	13 Jun 2018 11:02:01	U	SMPL
CRA - 1	CRA	966	0.1487	-	7470A	PPB	13 Jun 2018 11:04:08	U	SMPL
HighStd - 1	HighStd	38143	9.9700	-	7470A	PPB	13 Jun 2018 11:06:14	U	SMPL
ChkStd - 1	ChkStd	27325	7.1121	-	7470A	PPB	13 Jun 2018 11:08:21	U	SMPL
PB110173BL - 1	PBW	262	-0.0372	-	7470A	PPB	13 Jun 2018 11:13:03	U	SMPL
PB110173BS - 1	LCSW	13792	3.5371	-	7470A	PPB	13 Jun 2018 11:15:10	U	SMPL
PB110173TB - 1	PB110145TB	133	-0.0713	-	7470A	PPB	13 Jun 2018 11:17:17	U	SMPL
J2979-03 - 1	STONE-DUST-CANVAS-BAGS	240	-0.0430	-	7470A	PPB	13 Jun 2018 11:19:26	U	SMPL
J3245-02 - 1	CL-01-061118-A	295	-0.0285	-	7470A	PPB	13 Jun 2018 11:21:34	U	SMPL
J3245-04 - 1	CL-01-061118-B	252	-0.0399	-	7470A	PPB	13 Jun 2018 11:23:43	U	SMPL
J3245-06 - 1	CL-01-061118-C	302	-0.0267	-	7470A	PPB	13 Jun 2018 11:25:49	U	SMPL
CCV17 - 1	CCV17	19321	4.9977	-	7470A	PPB	13 Jun 2018 11:27:55	U	SMPL
CCB17 - 1	CCB17	215	-0.0497	-	7470A	PPB	13 Jun 2018 11:30:02	U	SMPL
J3245-08 - 1	CL-01-061118-D	512	0.0288	-	7470A	PPB	13 Jun 2018 11:32:09	U	SMPL
J3247-04 - 1	OR-03-061118	194	-0.0552	-	7470A	PPB	13 Jun 2018 11:34:16	U	SMPL
J3429-03 - 1	OILY-STONES	367	-0.0095	-	7470A	PPB	13 Jun 2018 11:36:23	U	SMPL
J3429-03DUP - 1	OILY-STONESDUP	342	-0.0161	-	7470A	PPB	13 Jun 2018 11:38:30	U	SMPL
J3429-03MS - 1	OILY-STONESMS	13891	3.5632	-	7470A	PPB	13 Jun 2018 11:40:37	U	SMPL
J3429-03MSD - 1	OILY-STONESMSD	13370	3.4256	-	7470A	PPB	13 Jun 2018 11:42:45	U	SMPL
J3442-02 - 1	CAS1-WC-S-01-1-S7	28	-0.0991	-	7470A	PPB	13 Jun 2018 11:44:53	U	SMPL
J3307-01 - 1	IDW-SO-06042018	936	0.1408	-	7470A	PPB	13 Jun 2018 11:47:01	U	SMPL
PB110174BL - 1	PBW	166	-0.0626	-	7470A	PPB	13 Jun 2018 11:49:10	U	SMPL
PB110174BS - 1	LCSW	13268	3.3986	-	7470A	PPB	13 Jun 2018 11:51:17	U	SMPL
CCV18 - 1	CCV18	19480	5.0397	-	7470A	PPB	13 Jun 2018 11:53:24	U	SMPL
CCB18 - 1	CCB18	198	-0.0541	-	7470A	PPB	13 Jun 2018 11:55:31	U	SMPL
PB110174TB - 1	PB110147TB	302	-0.0267	-	7470A	PPB	13 Jun 2018 11:57:39	U	SMPL
J3425-01 - 1	PGWPAL-AQ060818	684	0.0742	-	7470A	PPB	13 Jun 2018 12:04:22	U	SMPL
J3425-01DUP - 1	PGWPAL-AQ060818DUP	530	0.0336	-	7470A	PPB	13 Jun 2018 12:06:29	U	SMPL
J3425-01MS - 1	PGWPAL-AQ060818MS	13570	3.4784	-	7470A	PPB	13 Jun 2018 12:11:16	U	SMPL
J3425-01MSD - 1	PGWPAL-AQ060818MSD	13400	3.4335	-	7470A	PPB	13 Jun 2018 12:13:23	U	SMPL
J3429-03LX5 - 1		135	-0.0708	-	7470A	PPB	13 Jun 2018 12:15:31	U	SMPL
J3429-03A - 1		15534	3.9972	-	7470A	PPB	13 Jun 2018 12:17:39	U	SMPL
J3425-01LX5 - 1		169	-0.0618	-	7470A	PPB	13 Jun 2018 12:19:47	U	SMPL
J3425-01A - 1		14376	3.6913	-	7470A	PPB	13 Jun 2018 12:21:56	U	SMPL
CCV19 - 1	CCV19	19214	4.9694	-	7470A	PPB	13 Jun 2018 12:24:03	U	SMPL
CCB19 - 1	CCB19	133	-0.0713	-	7470A	PPB	13 Jun 2018 12:26:13	U	SMPL