



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB95980

Ambient Barometric Pressure (mmHg): 755

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	80.9	06/13/2018	09:40
2	J3448-02		1	>100.0	>212.0	>212.0	06/13/2018	10:05
3	J3448-02DUP		1	>100.0	>212.0	>212.0	06/13/2018	10:40
4	J3448-04		1	>100.0	>212.0	>212.0	06/13/2018	11:10
5	J3448-06		1	>100.0	>212.0	>212.0	06/13/2018	11:45
6	J3451-02		1	>100.0	>212.0	>212.0	06/13/2018	12:15
7	J3451-04		1	>100.0	>212.0	>212.0	06/13/2018	12:50
8	J3456-02		1	>100.0	>212.0	>212.0	06/13/2018	13:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 95980

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP61318

WorkList ID : 113276

Date : 6/13/2018 9:06:49 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2018	Water	J3448-02	C Flash Point	5°C Cool 4 deg C	PSEG01	D11	N-48971	06/12/2018	1010A
06/15/2018	Water	J3448-04	C Flash Point	7°C Cool 4 deg C	PSEG01	D11	N-48969	06/12/2018	1010A
06/15/2018	Water	J3448-06	C Flash Point	7°C Cool 4 deg C	PSEG01	D11	N-48966	06/12/2018	1010A
06/15/2018	Water	J3451-02	C Flash Point	6°C Cool 4 deg C	PSEG01	D11	VB-25830-SLUDGEBOX	06/12/2018	1010A
06/15/2018	Water	J3451-04	C Flash Point	5°C Cool 4 deg C	PSEG01	D11	VB-44949-SLUDGEBOX	06/12/2018	1010A
06/15/2018	Water	J3456-02	C Flash Point	6°C Cool 4 deg C	PSEG01	D12	ES-OILY-WATER	06/12/2018	1010A

Date/Time 06/13/18 9:10 AM

Received by: RS

Relinquished by: CP

Date/Time 06/13/18 13:25

Received by: CP

Relinquished by: KS