


7471B

LB95984
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47180



A= 0.0000e+000

B= 2.1386e-004

C= -3.3283e-002

Rho= 0.9997701

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.035	0.035	318	0.000	318				
0.2	0.200	0.187	-0.013	1029	0.0 %	1029				
2.5	2.500	2.448	-0.052	11604	0.0 %	11604				
5.0	5.000	5.109	0.109	24047	0.0 %	24047				
7.5	7.500	7.364	-0.136	34592	0.0 %	34592				
10.0	10.000	10.056	0.056	47180	0.0 %	47180				

LB95984 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Jun 2018 16:16:52

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	318	-	0.00007471B	PPB		13 Jun 2018 16:19:51	S	Std
0.2 - 1	SO.2	1029	-	0.20007471B	PPB		13 Jun 2018 16:22:09	S	Std
2.5 - 1	52.5	11604	-	2.50007471B	PPB		13 Jun 2018 16:24:27	S	Std
5.0 - 1	55	24047	-	5.00007471B	PPB		13 Jun 2018 16:26:44	S	Std
7.5 - 1	52.5	34592	-	7.50007471B	PPB		13 Jun 2018 16:29:02	S	Std
10.0 - 1	52.5	47180	-	10.00007471B	PPB		13 Jun 2018 16:34:28	S	Std
ICV99 - 1	ICV99	18156	3.8495	-7471B	PPB		13 Jun 2018 16:37:26	U	SMPL
ICB99 - 1	ICB99	173	0.0037	-7471B	PPB		13 Jun 2018 16:39:41	U	SMPL
CCV20 - 1	CCV20	24170	5.1356	-7471B	PPB		13 Jun 2018 16:41:57	U	SMPL
CCB20 - 1	CCB20	184	0.0061	-7471B	PPB		13 Jun 2018 16:44:13	U	SMPL
CRA - 1	CRA	1013	0.1834	-7471B	PPB		13 Jun 2018 16:46:29	U	SMPL
HighStd - 1	HighStd	46867	9.9895	-7471B	PPB		13 Jun 2018 16:48:45	U	SMPL
ChkStd - 1	ChkStd	33100	7.0454	-7471B	PPB		13 Jun 2018 16:51:01	U	SMPL
PB110194BL - 1	PBS	164	0.0018	-7471B	PPB		13 Jun 2018 16:55:53	U	SMPL
PB110194BS - 1	LCSS	15814	3.3486	-7471B	PPB		13 Jun 2018 16:58:09	U	SMPL
J3246-01 - 1	OK-02-061218	16492	3.4936	-7471B	PPB		13 Jun 2018 17:11:16	U	SMPL
J3246-01DUP - 1	OK-02-061218DUP	16698	3.5377	-7471B	PPB		13 Jun 2018 17:13:33	U	SMPL
J3246-01MS - 1	OK-02-061218MS	33014	7.0270	-7471B	PPB		13 Jun 2018 17:18:24	U	SMPL
J3246-01MSD - 1	OK-02-061218MSD	33278	7.0834	-7471B	PPB		13 Jun 2018 17:20:41	U	SMPL
J3251-03 - 1	SU-03-061118	1830	0.3581	-7471B	PPB		13 Jun 2018 17:22:58	U	SMPL
CCV21 - 1	CCV21	24090	5.1185	-7471B	PPB		13 Jun 2018 17:25:15	U	SMPL
CCB21 - 1	CCB21	152	-0.0008	-7471B	PPB		13 Jun 2018 17:27:31	U	SMPL
J3458-01 - 1	EPE-105-5(0-5)	6136	1.2789	-7471B	PPB		13 Jun 2018 17:29:47	U	SMPL
J3458-02 - 1	EPE-105-5(5-10)	1949	0.3835	-7471B	PPB		13 Jun 2018 17:32:03	U	SMPL
J3458-03 - 1	EPE-105-5(10-12)	1639	0.3172	-7471B	PPB		13 Jun 2018 17:34:19	U	SMPL
J3458-04 - 1	EPE-105-5(45-47)	802	0.1382	-7471B	PPB		13 Jun 2018 17:36:36	U	SMPL
J3246-01LX5 - 1		3597	0.7360	-7471B	PPB		13 Jun 2018 17:38:53	U	SMPL
J3246-01A - 1		31617	6.7282	-7471B	PPB		13 Jun 2018 17:41:10	U	SMPL
CCV22 - 1	CCV22	24087	5.1179	-7471B	PPB		13 Jun 2018 17:43:27	U	SMPL
CCB22 - 1	CCB22	164	0.0018	-7471B	PPB		13 Jun 2018 17:45:46	U	SMPL