

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/14/2018

TEMP1 IN: 103 06/14/2018 07:15 TEMP OUT: 104 06/14/2018 08:20

TEMP2 IN: 104 06/14/2018 09:00 TEMP OUT: 104 06/14/2018 10:05

TEMP3 IN: 105 06/14/2018 12:00 TEMP OUT: 105 06/14/2018 13:34

TEMP4 IN: 105 06/14/2018 14:20 TEMP OUT: 105 06/14/2018 16:10

Run Number: LB95990

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB95990BL	LB95990BL	1.3948	1.3948	100	1.3948	1.3948	1.3948	0.0000	0
2	LB95990BS	LB95990BS	1.3974	1.3974	100	1.4470	1.4470	1.4470	0.0496	496
3	J3466-01	COMP	1.3889	1.3889	100	1.3977	1.3977	1.3977	0.0088	88
4	J3469-01	OILY-WATER	1.3947	1.3947	50	1.4580	1.4580	1.4580	0.0633	1266
5	J3469-01DUP	OILY-WATERDUP	1.3956	1.3956	50	1.4570	1.4570	1.4570	0.0614	1228
6	J3474-01	061118-DBRI-E-D-B	1.3905	1.3905	400	1.4073	1.4073	1.4073	0.0168	42
7	J3474-02	061118-DBRI-E-D-M	1.3935	1.3935	400	1.4074	1.4074	1.4074	0.0139	34.8
8	J3474-03	061118-DBRI-E-D-S	1.3929	1.3929	400	1.4023	1.4023	1.4023	0.0094	23.5
9	J3474-04	061118-DBRI-E-U-B	1.4015	1.4015	400	1.4165	1.4165	1.4165	0.0150	37.5
10	J3474-05	061118-DBRI-E-U-M	1.3970	1.3970	400	1.4078	1.4078	1.4078	0.0108	27
11	J3474-06	061118-DBRI-E-U-S	1.4001	1.4001	400	1.4087	1.4087	1.4087	0.0086	21.5
12	J3475-01	061218-DBRI-F-D-B	1.4103	1.4103	400	1.4277	1.4277	1.4277	0.0174	43.5
13	J3475-02	061218-DBRI-F-D-M	1.4105	1.4105	400	1.4235	1.4235	1.4235	0.0130	32.5
14	J3475-03	061218-DBRI-F-D-S	1.3956	1.3956	400	1.4069	1.4069	1.4069	0.0113	28.3
15	J3475-04	061218-DBRI-F-D-B	1.3950	1.3950	400	1.4138	1.4138	1.4138	0.0188	47

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/14/2018

TEMP1 IN: 103 06/14/2018 07:15 TEMP OUT: 104 06/14/2018 08:20

TEMP2 IN: 104 06/14/2018 09:00 TEMP OUT: 104 06/14/2018 10:05

TEMP3 IN: 105 06/14/2018 12:00 TEMP OUT: 105 06/14/2018 13:34

TEMP4 IN: 105 06/14/2018 14:20 TEMP OUT: 105 06/14/2018 16:10

Run Number: LB95990

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J3475-05	061218-DBRI-F-D-M	1.3946	1.3946	400	1.4094	1.4094	1.4094	0.0148	37
17	J3475-06	061218-DBRI-F-D-S	1.4037	1.4037	400	1.4153	1.4153	1.4153	0.0116	29
18	J3495-01	DSN-002A-061318	1.3888	1.3888	600	1.4035	1.4035	1.4035	0.0147	24.5
19	J3503-01	WS-6	1.3985	1.3985	200	1.4110	1.4110	1.4110	0.0125	62.5
20	J3503-01DUP	WS-6DUP	1.3943	1.3943	200	1.4070	1.4070	1.4070	0.0127	63.5
21	J3503-02	WS-4	1.4080	1.4080	20	1.4289	1.4289	1.4289	0.0209	1045
22	J3503-03	WS-5	1.4036	1.4036	100	1.4067	1.4067	1.4067	0.0031	31

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061418

WorkList ID : 113310

Date : 6/14/2018 7:24:19 AM

2195990

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Water	J3466-01 <i>ND</i>	TSS	Cool 4 deg C	ARAM01	D41	COMP	06/13/2018	SM2540 D
06/17/2018	Water	J3469-01 <i>F</i>	TSS	Cool 4 deg C	PSEG01	D21	OILY-WATER	06/13/2018	SM2540 D
06/18/2018	Water	J3474-01 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-D-B	06/11/2018	SM2540 D
06/18/2018	Water	J3474-02 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-D-M	06/11/2018	SM2540 D
06/18/2018	Water	J3474-03 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-D-S	06/11/2018	SM2540 D
06/18/2018	Water	J3474-04 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-U-B	06/11/2018	SM2540 D
06/18/2018	Water	J3474-05 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-U-M	06/11/2018	SM2540 D
06/18/2018	Water	J3474-06 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061118-DBRI-E-U-S	06/11/2018	SM2540 D
06/19/2018	Water	J3475-01 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-B	06/12/2018	SM2540 D
06/19/2018	Water	J3475-02 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-M	06/12/2018	SM2540 D
06/19/2018	Water	J3475-03 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-S	06/12/2018	SM2540 D
06/19/2018	Water	J3475-04 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-B	06/12/2018	SM2540 D
06/19/2018	Water	J3475-05 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-M	06/12/2018	SM2540 D
06/19/2018	Water	J3475-06 <i>C</i>	TSS	Cool 4 deg C	HDRIO3	D11	061218-DBRI-F-D-S	06/12/2018	SM2540 D
06/20/2018	Water	J3495-01 <i>B</i>	TSS	Cool 4 deg C	PKFM01	D42	DSN-002A-061318	06/13/2018	SM2540 D
06/20/2018	Water	J3503-01 <i>F</i>	TSS	Cool 4 deg C	TULL02		WS-6	06/13/2018	SM2540 D
06/20/2018	Water	J3503-02 <i>F</i>	TSS	Cool 4 deg C	TULL02		WS-4	06/13/2018	SM2540 D
06/20/2018	Water	J3503-03 <i>F</i>	TSS	Cool 4 deg C	TULL02		WS-5	06/13/2018	SM2540 D

Date/Time

06/14/18 08:00

Received by:

AS

Relinquished by:

CP

Date/Time

06/14/18 13:00

Received by:

AS

Relinquished by:

AS