



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95994

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	06/13/2018	16:20
2	CAL2	1	Water			21.6	7.00	06/13/2018	16:22
3	CAL3	1	Water			21.6	10.03	06/13/2018	16:24
4	ICV	1	Water			21.8	7.00	06/13/2018	16:26
5	CCV1	1	Water			21.7	2.00	06/13/2018	16:28
6	J3469-01	1	Water			23.8	4.93	06/13/2018	16:30
7	J3469-01DUP	1	Water			23.9	4.94	06/13/2018	16:32
8	J3497-01	1	Water			22.0	8.29	06/13/2018	16:38
9	CCV2	1	Water			21.6	12.00	06/13/2018	16:40

WORKLIST(Hardcopy Internal Chain)

LB95994

WorkList Name : pH061318

WorkList ID : 113304

Date : 6/13/2018 3:50:09 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/13/2018	Water	J3469-01	pH	Cool 4 deg C	PSEG01	D21	OILY-WATER	06/13/2018	9040C
06/13/2018	Water	J3497-01	pH	Cool 4 deg C	PARS15	D42	CAS1-WC-L-01	06/13/2018	9040C

Date/Time 06/13/18 4:15 PM

Received by: SP

Relinquished by: CP

Date/Time 06/13/18 7:00 PM

Received by: CP

Relinquished by: SP