



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB95996

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.8	4.01	06/13/2018	16:20
2	CAL2	1	Water			21.6	7.00	06/13/2018	16:22
3	CAL3	1	Water			21.6	10.03	06/13/2018	16:24
4	ICV	1	Water			21.8	7.00	06/13/2018	16:26
5	CCV1	1	Water			21.8	2.00	06/13/2018	18:30
6	J3482-01	1	Solid	20.02	20	25.2	5.64	06/13/2018	18:32
7	J3482-01DUP	1	Solid	20.04	20	25.3	5.65	06/13/2018	18:34
8	CCV2	1	Water			21.7	12.00	06/13/2018	18:36

LB 95996

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH061318-1

WorkList ID : 113305

Date : 6/13/2018 3:50:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/12/2018	Solid	J3482-01	pH	Cool 4 deg C	TULL02	D61	FK-PS-01-029	06/12/2018	9045C

Date/Time 06/13/18 5:30 PM
Received by: SP
Relinquished by: CP

Date/Time 06/13/18 7:00 PM
Received by: CP
Relinquished by: SP