



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB95997

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	06/14/2018	09:00
2	CAL2	1	Water			21.8	7.00	06/14/2018	09:02
3	CAL3	1	Water			21.7	10.03	06/14/2018	09:04
4	ICV	1	Water			21.5	7.00	06/14/2018	09:06
5	CCV1	1	Water			21.5	2.00	06/14/2018	10:25
6	J3464-07	1	Solid	20.02	20	24.3	7.46	06/14/2018	10:27
7	J3464-07DUP	1	Solid	20.03	20	24.4	7.47	06/14/2018	10:29
8	J3464-08	1	Solid	20.02	20	24.2	5.82	06/14/2018	10:31
9	J3464-09	1	Solid	20.03	20	24.1	6.63	06/14/2018	10:33
10	J3464-10	1	Solid	20.02	20	24.1	6.90	06/14/2018	10:35
11	J3464-11	1	Solid	20.04	20	23.6	6.84	06/14/2018	10:37
12	J3464-12	1	Solid	20.02	20	24.0	6.62	06/14/2018	10:39
13	CCV2	1	Water			21.6	12.00	06/14/2018	10:41

WORKLIST(Hardcopy Internal Chain)

LB95997

WorkList Name : corrosivity061418

WorkList ID : 113350

Date : 6/14/2018 10:11:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/08/2018	Solid	J3464-07	Corrosivity	Cool 4 deg C	TULL02	D11	SB-38-COMP	06/13/2018	9045C
06/08/2018	Solid	J3464-08	Corrosivity	Cool 4 deg C	TULL02	D11	SB-39-COMP	06/13/2018	9045C
06/08/2018	Solid	J3464-09	Corrosivity	Cool 4 deg C	TULL02	D11	SB-40-COMP	06/13/2018	9045C
06/08/2018	Solid	J3464-10	Corrosivity	Cool 4 deg C	TULL02	D11	SB-41-COMP	06/13/2018	9045C
06/08/2018	Solid	J3464-11	Corrosivity	Cool 4 deg C	TULL02	D11	SB-42-COMP	06/13/2018	9045C
06/08/2018	Solid	J3464-12	Corrosivity	Cool 4 deg C	TULL02	D11	SB-43-COMP	06/13/2018	9045C

Date/Time 06/14/18 06:15 AM

Received by: Sp

Relinquished by: Sp

Date/Time 06/14/18 11:00 AM

Received by: Sp

Relinquished by: Sp