



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95999

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP65279
COD calibration std. 100 ppm	WP65276
COD calibration std. 10 ppm	WP65275
COD calibration std. 150 ppm	WP65277
COD calibration std. 50 ppm	WP65278
COD calibration std. 0 ppm	WP65274
COD CCV std, 50ppm	WP65280
COD Digestion Vials Low Level 0-150Mg/L	W2346

Temp In (C): <u>150</u>	Date In: <u>06/14/2018</u>	Time In: <u>09:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>06/14/2018</u>	Time Out: <u>11:02</u>

Intercept: 0.3627

Slope: 1.0071

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	06/14/2018	12:00
2	CAL2	10	1	Water	11.000	06/14/2018	12:00
3	CAL3	50	1	Water	50.000	06/14/2018	12:01
4	CAL4	100	1	Water	102.000	06/14/2018	12:01
5	CAL5	150	1	Water	151.000	06/14/2018	12:02



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB95999

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.287	06/14/2018	12:02
2	ICB				1	Water	0.000	-0.360	06/14/2018	12:03
3	CCV1	50			1	Water	51.000	50.280	06/14/2018	12:03
4	CCB1				1	Water	0.000	-0.360	06/14/2018	12:04
5	LB95999BL				1	Water	0.000	-0.360	06/14/2018	12:04
6	LB95999BS	50			1	Water	50.000	49.287	06/14/2018	12:05
7	J3448-01				1	Water	29.000	28.435	06/14/2018	12:05
8	J3448-01DUP				1	Water	31.000	30.421	06/14/2018	12:06
9	J3448-01MS	50			1	Water	70.000	69.146	06/14/2018	12:06
10	J3448-01MSD	50			1	Water	71.000	70.139	06/14/2018	12:07
11	J3448-03				1	Water	35.000	34.393	06/14/2018	12:07
12	J3448-05				10	Water	22.000	21.485	06/14/2018	12:08
13	J3451-01				1	Water	128.000	126.737	06/14/2018	12:08
14	J3451-03				1	Water	140.000	138.653	06/14/2018	12:09
15	CCV2	50			1	Water	51.000	50.280	06/14/2018	12:09
16	CCB2				1	Water	0.000	-0.360	06/14/2018	12:10
17	J3456-01				10	Water	70.000	69.146	06/14/2018	12:10
18	CCV3	50			1	Water	51.000	50.280	06/14/2018	12:11
19	CCB3				1	Water	0.000	-0.360	06/14/2018	12:11