

OVEN TEMP1: 103 °C TEMP: 105 °C

DATE & TIME IN: 06/14/2018 15:00 OUT: 06/15/2018 07:20

OVEN TEMP2: 105 °C TEMP: 105 °C

DATE & TIME IN: 06/15/2018 08:00 OUT: 06/15/2018 09:30

Run Number: LB96001

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB96001BL	LB96001BL	95.3635	95.3635	100	95.3635	95.3635	95.3635	0.0000	0
2	J3503-01	WS-6	93.0351	93.0351	100	93.1656	93.1656	93.1656	0.1305	1305
3	J3503-02	WS-4	82.6383	82.6383	100	82.9160	82.9160	82.9160	0.2777	2777
4	J3503-02DUP	WS-4DUP	89.1641	89.1641	100	89.4299	89.4299	89.4299	0.2658	2658
5	J3503-03	WS-5	84.6078	84.6078	100	84.6261	84.6261	84.6261	0.0183	183

A = Sample Volume (ml)
 B = Empty Dish Weight (g)
 C = Final Empty Dish + Sample weight after drying @105°C(g)
 D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ts061418

WorkList ID : 113366

Date : 6/14/2018 12:22:43 PM

1396001

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Water	J3503-01	TS	Cool 4 deg C	TULL02		WS-6	06/13/2018	SM2540 B
06/20/2018	Water	J3503-02	TS	Cool 4 deg C	TULL02		WS-4	06/13/2018	SM2540 B
06/20/2018	Water	J3503-03	TS	Cool 4 deg C	TULL02		WS-5	06/13/2018	SM2540 B

Date/Time 6/14/18, 14:00
 Received by: AS
 Relinquished by: CP

Date/Time 6/14/18, 16:00
 Received by: CP
 Relinquished by: AS