



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96019

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	22	27.0	80.6	80.9	06/14/2018	09:45
2	J3469-02		1	7	>100.0	>212.0	>212.0	06/14/2018	10:10
3	J3469-02DUP		1	7	>100.0	>212.0	>212.0	06/14/2018	10:35
4	J3485-08		1	5	>100.0	>212.0	>212.0	06/14/2018	10:55
5	J3497-02		1	6	>100.0	>212.0	>212.0	06/14/2018	11:15
6	J3503-01		1	4	>100.0	>212.0	>212.0	06/14/2018	11:40
7	J3503-02		1	8	>100.0	>212.0	>212.0	06/14/2018	12:02
8	J3503-03		1	7	>100.0	>212.0	>212.0	06/14/2018	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 96019

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP61418

WorkList ID : 113338

Date : 6/14/2018 9:04:52 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2018	Water	J3469-02 C	Flash Point 7°C	Cool 4 deg C	PSEG01	D21	OILY-WATER	06/13/2018	1010A
06/26/2018	Water	J3485-08 D	Flash Point 5°C	Cool 4 deg C	HALE01	D52	IDW-GW-06132018	06/13/2018	1010A
06/26/2018	Water	J3497-02 B	Flash Point 6°C	Cool 4 deg C	PARS15	D42	CAS1-WC-L-01	06/13/2018	1010A
06/20/2018	Water	J3503-01 K	Flash Point 4°C	Cool 4 deg C	TULL02		WS-6	06/13/2018	1010A
06/20/2018	Water	J3503-02 K	Flash Point 8°C	Cool 4 deg C	TULL02		WS-4	06/13/2018	1010A
06/20/2018	Water	J3503-03 K	Flash Point 7°C	Cool 4 deg C	TULL02		WS-5	06/13/2018	1010A

Date/Time 06/14/18 9:15 AM

Received by: KS

Relinquished by: AP

Date/Time 06/14/18 12:30 PM

Received by: AP

Relinquished by: KS