



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96022

Slope : 98.1

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/14/2018	16:00
2	CAL2	1	Water			21.8	7.00	06/14/2018	16:02
3	CAL3	1	Water			21.5	10.03	06/14/2018	16:04
4	ICV	1	Water			21.6	7.00	06/14/2018	16:06
5	CCV1	1	Water			21.5	2.00	06/14/2018	16:08
6	J3508-02	1	Water			21.7	7.05	06/14/2018	16:10
7	J3508-02DUP	1	Water			21.8	7.06	06/14/2018	16:12
8	J3508-05	1	Water			21.6	7.33	06/14/2018	16:14
9	J3512-02	1	Water			22.5	7.44	06/14/2018	16:16
10	CCV2	1	Water			21.5	12.00	06/14/2018	16:18

LB 96022

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity061418-1

WorkList ID : 113377

Date : 6/14/2018 3:43:43 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2018	Water	J3508-02	Corrosivity	Cool 4 deg C	PSEG01	D31	KGS-SIDE-1-A	06/14/2018	9040C
06/14/2018	Water	J3508-05	Corrosivity	Cool 4 deg C	PSEG01	D31	KGS-SIDE-2-3-A	06/14/2018	9040C
06/14/2018	Water	J3512-02	Corrosivity	Cool 4 deg C	PSEG01	D32	FADW4710L	06/14/2018	9040C

Date/Time 06/14/18 3:50 PM

Received by: SP

Relinquished by: CP

Date/Time 06/14/18 4:30 PM

Received by: CP

Relinquished by: SP