


LB 96032

245.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 3.0052e-004

C= -6.6685e-002

Rho= 0.9996623

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.029	-0.029	127	0.000	127				
0.05	0.050	0.061	0.011	425	0.0 %	425				
0.2	0.200	0.237	0.037	1012	0.0 %	1012				
2.5	2.500	2.554	0.054	8719	0.0 %	8719				
5.0	5.000	4.997	-0.003	16849	0.0 %	16849				
7.5	7.500	7.295	-0.205	24498	0.0 %	24498				
10.0	10.000	10.134	0.134	33945	0.0 %	33945				

LB96032 INSTRUMENT ID : CV1

Method: 245.1	Operator: Admin	Date of Analysis: 15 Jun 2018 09:06:12							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	90	127	-	0.0000	245.1	PPB	15 Jun 2018 09:27:12	S	Std
0.05 - 1	51.05	425	-	0.0500	245.1	PPB	15 Jun 2018 09:29:21	S	Std
0.2 - 1	50.2	1012	-	0.2000	245.1	PPB	15 Jun 2018 09:31:29	S	Std
2.5 - 1	52.5	8719	-	2.5000	245.1	PPB	15 Jun 2018 09:33:37	S	Std
5.0 - 1	55	16849	-	5.0000	245.1	PPB	15 Jun 2018 09:35:45	S	Std
7.5 - 1	57.5	24498	-	7.5000	245.1	PPB	15 Jun 2018 09:37:52	S	Std
10.0 - 1	57.5	33945	-	10.0000	245.1	PPB	15 Jun 2018 09:40:01	S	Std
ICV03 - 1	ICV03	13363	3.9491	-245.1		PPB	15 Jun 2018 09:45:51	U	SMPL
ICB03 - 1	ICB03	111	-0.0333	-245.1		PPB	15 Jun 2018 09:47:57	U	SMPL
CCV35 - 1	CCV35	16996	5.0409	-245.1		PPB	15 Jun 2018 09:50:04	U	SMPL
CCB35 - 1	CCB35	82	-0.0420	-245.1		PPB	15 Jun 2018 09:52:11	U	SMPL
CRA - 1	CRA	411	0.0568	-245.1		PPB	15 Jun 2018 09:56:50	U	SMPL
HighStd - 1	HighStd	33348	9.9550	-245.1		PPB	15 Jun 2018 09:58:56	U	SMPL
ChkStd - 1	ChkStd	23206	6.9071	-245.1		PPB	15 Jun 2018 10:01:03	U	SMPL
PB110242BL - 1	PBW	137	-0.0255	-245.1		PPB	15 Jun 2018 10:05:41	U	SMPL
PB110242BS - 1	LCSW	12696	3.7487	-245.1		PPB	15 Jun 2018 10:07:47	U	SMPL
J3474-01 - 1	061118-DBRI-E-D-B	-139	-0.1085	-245.1		PPB	15 Jun 2018 10:09:55	U	SMPL
J3474-02 - 1	061118-DBRI-E-D-M	-20	-0.0727	-245.1		PPB	15 Jun 2018 10:12:04	U	SMPL
J3474-03 - 1	061118-DBRI-E-D-S	-24	-0.0739	-245.1		PPB	15 Jun 2018 10:14:13	U	SMPL
J3474-04 - 1	061118-DBRI-E-U-B	4	-0.0655	-245.1		PPB	15 Jun 2018 10:16:22	U	SMPL
J3474-05 - 1	061118-DBRI-E-U-M	25	-0.0592	-245.1		PPB	15 Jun 2018 10:18:28	U	SMPL
CCV36 - 1	CCV36	17028	5.0505	-245.1		PPB	15 Jun 2018 10:20:34	U	SMPL
CCB36 - 1	CCB36	86	-0.0408	-245.1		PPB	15 Jun 2018 10:22:40	U	SMPL
J3474-06 - 1	061118-DBRI-E-U-S	-26	-0.0745	-245.1		PPB	15 Jun 2018 10:24:48	U	SMPL
J3475-01 - 1	061218-DBRI-F-D-B	17	-0.0616	-245.1		PPB	15 Jun 2018 10:26:55	U	SMPL
J3475-02 - 1	061218-DBRI-F-D-M	6	-0.0649	-245.1		PPB	15 Jun 2018 10:29:02	U	SMPL
J3475-03 - 1	061218-DBRI-F-D-S	1	-0.0664	-245.1		PPB	15 Jun 2018 10:31:09	U	SMPL
J3475-04 - 1	061218-DBRI-F-D-B	11	-0.0634	-245.1		PPB	15 Jun 2018 10:33:16	U	SMPL
J3475-05 - 1	061218-DBRI-F-D-M	-23	-0.0736	-245.1		PPB	15 Jun 2018 10:35:24	U	SMPL
J3475-06 - 1	061218-DBRI-F-D-S	36	-0.0559	-245.1		PPB	15 Jun 2018 10:37:32	U	SMPL
J3503-01 - 1	WS-6	83	-0.0417	-245.1		PPB	15 Jun 2018 10:39:40	U	SMPL
J3503-02 - 1	WS-4	1744	0.4574	-245.1		PPB	15 Jun 2018 10:41:49	U	SMPL
J3503-03 - 1	WS-5	249	0.0081	-245.1		PPB	15 Jun 2018 10:43:56	U	SMPL
CCV37 - 1	CCV37	16968	5.0325	-245.1		PPB	15 Jun 2018 10:46:02	U	SMPL
CCB37 - 1	CCB37	119	-0.0309	-245.1		PPB	15 Jun 2018 10:48:09	U	SMPL
J3503-03DUP - 1	WS-5DUP	236	0.0042	-245.1		PPB	15 Jun 2018 10:50:18	U	SMPL
J3503-03MS - 1	WS-5MS	14007	4.1427	-245.1		PPB	15 Jun 2018 10:52:25	U	SMPL
J3503-03MSD - 1	WS-5MSD	13010	3.8430	-245.1		PPB	15 Jun 2018 10:54:31	U	SMPL
J3503-03LX5 - 1		183	-0.0117	-245.1		PPB	15 Jun 2018 10:56:38	U	SMPL
J3503-03A - 1		14643	4.3338	-245.1		PPB	15 Jun 2018 10:58:46	U	SMPL
CCV38 - 1	CCV38	16984	5.0373	-245.1		PPB	15 Jun 2018 11:00:53	U	SMPL
CCB38 - 1	CCB38	128	-0.0282	-245.1		PPB	15 Jun 2018 11:03:02	U	SMPL