



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB96034
Analysis Date: 06/15/2018
BalanceID: WC SC-3
OvenID: EX Oven-2

ANALYST: Jignesh
REVIEWED BY: apatel
Extraction Date: 06/15/2018
Extraction IN Time: 07:45
Extraction OUT Time: 08:30

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB96034BL	LB96034BL	WATER	2	1000	100	2.9603	2.9603	3.02	2.9603	2.9603	0.0000	0
2	LB96034BS	LB96034BS	WATER	2	1000	100	3.0255	3.0255	3.03	3.0445	3.0445	0.0190	19
3	LB96034BSD	LB96034BSD	WATER	2	1000	100	3.0488	3.0488	3.04	3.0680	3.0680	0.0192	19.2
4	J3469-01	OILY-WATER	WATER	2	800	100	3.1475	3.1475	12.00	3.3263	3.3263	0.1788	223.5
5	J3508-01	KGS-SIDE-1	WATER	2	1000	100	3.0469	3.0469	3.01	3.0497	3.0497	0.0028	2.8
6	J3508-04	KGS-SIDE-2-3	WATER	2	1000	100	3.0643	3.0643	3.02	3.0664	3.0664	0.0021	2.1
7	J3512-01	FADW4710L	WATER	2	1000	100	3.0429	3.0429	9.00	3.1182	3.1182	0.0753	75.3



QC Batch# LB96034 Test: TPH Analysis Date: 06/15/2018

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2374
pH Paper 0-14	W2373
Sodium Sulfate	EP1853
1:1 HCL	WP64228
Silica Gel	W2199
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP64229
LCSWD	10.00 ML	WP64230
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP :	<u>70</u>	Dessicator Time In :	<u>10:11</u>
1.0000 gram Balance:	<u>1.0003</u>	In Time:	<u>09:00</u>		
Bal Check Time:	<u>08:10</u>	Out OVEN TEMP:	<u>70</u>	Dessicator Time Out:	<u>10:35</u>
		Out Time:	<u>10:10</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP :	<u>71</u>	Dessicator Time In :	<u>11:31</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time:	<u>11:00</u>		
Bal Check Time:	<u>12:35</u>	Out OVEN TEMP:	<u>71</u>	Dessicator Time Out:	<u>12:05</u>
		Out Time:	<u>11:30</u>		

2B 96034

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TPH J3508

WorkList ID : 113388

Date : 6/15/2018 7:37:26 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/17/2018	Water	J3469-01 -K	TPH	Conc H2SO4 4c	PSEG01	D21	OILY-WATER	06/13/2018	1664A
06/19/2018	Water	J3508-01 -L	TPH	Conc H2SO4 4c	PSEG01	D31	KGS-SIDE-1	06/14/2018	1664A
06/19/2018	Water	J3508-04 -L	TPH	Conc H2SO4 4c	PSEG01	D31	KGS-SIDE-2-3	06/14/2018	1664A
06/19/2018	Water	J3512-01 -L	TPH	Conc H2SO4 4c	PSEG01	D32	FADW4710L	06/14/2018	1664A

Date/Time 06-15-18 7:46 AM
 Received by: JP
 Relinquished by: CS

Date/Time 06-15-18 10:20 AM
 Received by: CS
 Relinquished by: JP