



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96039

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP65359
Turbidity std, 40NTU	WP65354
turbidity std 5 NTU	WP65357
turbidity std 20 NTU	WP65355
Turbidity std 1 NTU	WP65358
turbidity std 0 NTU	WP65353
turbidity std 10 NTU	WP65356
10 NTU Standard 500 ml	W2242

Intercept: -0.0236

Slope: 0.9158

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.180	06/15/2018	12:49
2	CAL2	1	1	1.020	06/15/2018	12:52
3	CAL3	5	1	4.680	06/15/2018	12:57
4	CAL4	10	1	8.880	06/15/2018	13:02
5	CAL5	20	1	17.800	06/15/2018	13:06
6	CAL6	40	1	36.900	06/15/2018	13:10



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96039

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.610	10.519	06/15/2018	13:14
2	ICB		1	0.182	0.225	06/15/2018	13:19
3	CCV1	10	1	9.160	10.028	06/15/2018	13:23
4	CCB1		1	0.195	0.239	06/15/2018	13:27
5	LB96039BL		1	0.186	0.229	06/15/2018	13:30
6	J3547-01		1	0.242	0.290	06/15/2018	13:34
7	J3547-01DUP		1	0.240	0.288	06/15/2018	13:37
8	CCV2	10	1	9.280	10.159	06/15/2018	13:41
9	CCB2		1	0.182	0.225	06/15/2018	13:44

CB 96039

WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity061518

WorkList ID : 113425

Date : 6/15/2018 11:52:31 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2018	Water	J3547-01	Turbidity	Cool 4 deg C	LIRO01	D42	UDS-MF-01	06/13/2018	SM2130 B

Date/Time 6-15-18 12:18
 Received by: AL
 Relinquished by: CA

Date/Time 6-15-18 14:10
 Received by: CA
 Relinquished by: AL