

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/15/2018

TEMP1 IN: 103 06/15/2018 07:10 TEMP OUT: 104 06/15/2018 08:15

TEMP2 IN: 104 06/15/2018 09:00 TEMP OUT: 104 06/15/2018 10:10

TEMP3 IN: 105 06/15/2018 12:30 TEMP OUT: 105 06/15/2018 14:05

TEMP4 IN: 105 06/15/2018 14:45 TEMP OUT: 105 06/15/2018 16:25

Run Number: LB96043

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96043BL	LB96043BL	1.3895	1.3895	100	1.3895	1.3895	1.3895	0.0000	0
2	LB96043BS	LB96043BS	1.3541	1.3541	100	1.4050	1.4050	1.4050	0.0509	509
3	J3547-01	UDS-MF-01	1.3980	1.3980	400	1.4085	1.4085	1.4085	0.0105	26.3
4	J3547-01DUP	UDS-MF-01DUP	1.4004	1.4004	400	1.4105	1.4105	1.4105	0.0101	25.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061518-2

WorkList ID : 113424

Date : 6/15/2018 11:51:59 AM

DB96643

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2018	Water	J3547-01	TSS	Cool 4 deg C	LIR001	D42	UDS-MF-01	06/13/2018	SM2540 D

ND

Date/Time

06/15/18, 12:00

Received by:

AS

Relinquished by:

AS

Date/Time

06/15/18, 13:00

Received by:

AS

Relinquished by:

AS