


7471B

LB 96046
INSTRUMENT ID : CV1

Linear

 μ Abs.:

42904



A= 0.0000e+000

B= 2.3329e-004

C= -3.4010e-002

Rho= 0.9999590

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.044	0.044	334	0.000	334				
0.2	0.200	0.172	-0.028	883	0.0 %	883				
2.5	2.500	2.466	-0.034	10716	0.0 %	10716				
5.0	5.000	4.996	-0.004	21561	0.0 %	21561				
7.5	7.500	7.547	0.047	32495	0.0 %	32495				
10.0	10.000	9.975	-0.025	42904	0.0 %	42904				

LB96046 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 15 Jun 2018 15:06:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	334	-	0.00007471B	PPB		15 Jun 2018 15:23:29	S	Std
0.2 - 1	SO-2	883	-	0.20007471B	PPB		15 Jun 2018 15:25:46	S	Std
2.5 - 1	S2-T	10716	-	2.50007471B	PPB		15 Jun 2018 15:28:04	S	Std
5.0 - 1	S5	21561	-	5.00007471B	PPB		15 Jun 2018 15:30:21	S	Std
7.5 - 1	S7.5	32495	-	7.50007471B	PPB		15 Jun 2018 15:32:38	S	Std
10.0 - 1	S10	42904	-	10.00007471B	PPB		15 Jun 2018 15:37:35	S	Std
ICV05 - 1	ICV05	17154	3.9679	-7471B	PPB		15 Jun 2018 15:40:14	U	SMPL
ICB05 - 1	ICB05	-114	-0.0606	-7471B	PPB		15 Jun 2018 15:42:31	U	SMPL
CCV45 - 1	CCV45	21359	4.9489	-7471B	PPB		15 Jun 2018 15:44:47	U	SMPL
CCB45 - 1	CCB45	-116	-0.0611	-7471B	PPB		15 Jun 2018 15:47:02	U	SMPL
CRA - 1	CRA	902	0.1764	-7471B	PPB		15 Jun 2018 15:49:19	U	SMPL
HighStd - 1	HighStd	43017	10.0016	-7471B	PPB		15 Jun 2018 15:51:35	U	SMPL
ChkStd - 1	ChkStd	31990	7.4291	-7471B	PPB		15 Jun 2018 15:53:51	U	SMPL
PB110276BL - 1	PBS	-137	-0.0660	-7471B	PPB		15 Jun 2018 15:56:10	U	SMPL
PB110276BS - 1	LCSS	16253	3.7577	-7471B	PPB		15 Jun 2018 15:58:28	U	SMPL
J3464-01 - 1	SB-38-COMP	1798	0.3855	-7471B	PPB		15 Jun 2018 16:00:45	U	SMPL
J3464-02 - 1	SB-39-COMP	302	0.0364	-7471B	PPB		15 Jun 2018 16:03:03	U	SMPL
J3464-03 - 1	SB-40-COMP	1569	0.3320	-7471B	PPB		15 Jun 2018 16:05:21	U	SMPL
J3464-04 - 1	SB-41-COMP	627	0.1123	-7471B	PPB		15 Jun 2018 16:07:39	U	SMPL
J3464-05 - 1	SB-42-COMP	924	0.1816	-7471B	PPB		15 Jun 2018 16:09:55	U	SMPL
CCV46 - 1	CCV46	21269	4.9279	-7471B	PPB		15 Jun 2018 16:12:12	U	SMPL
CCB46 - 1	CCB46	-109	-0.0594	-7471B	PPB		15 Jun 2018 16:14:28	U	SMPL
J3464-06 - 1	SB-43-COMP	666	0.1214	-7471B	PPB		15 Jun 2018 16:16:44	U	SMPL
J3486-01 - 1	DOCUMENTATION-4	20562	4.7630	-7471B	PPB		15 Jun 2018 16:19:00	U	SMPL
J3486-01DUP - 1	DOCUMENTATION-4DUP	20476	4.7429	-7471B	PPB		15 Jun 2018 16:26:37	U	SMPL
J3486-01MS - 1	DOCUMENTATION-4MS	35795	8.3167	-7471B	PPB		15 Jun 2018 16:28:53	U	SMPL
J3486-01MSD - 1	DOCUMENTATION-4MSD	35554	8.2605	-7471B	PPB		15 Jun 2018 16:31:09	U	SMPL
J3486-02 - 1	DOCUMENTATION-5	23531	5.4556	-7471B	PPB		15 Jun 2018 16:33:28	U	SMPL
J3486-03 - 1	DOCUMENTATION-6	51472	11.9741	-7471B	PPB		15 Jun 2018 16:35:47	U	SMPL
J3486-04 - 1	DOCUMENTATION-7	292530	HIGH	-7471B	PPB		15 Jun 2018 16:38:05	X	
J3486-05 - 1	DOCUMENTATION-8	43933	10.2153	-7471B	PPB		15 Jun 2018 16:40:01	U	SMPL
J3486-06 - 1	DOCUMENTATION-9	77865	18.1314	-7471B	PPB		15 Jun 2018 16:42:22	U	SMPL
CCV47 - 1	CCV47	20822	4.8236	-7471B	PPB		15 Jun 2018 16:44:40	U	SMPL
CCB47 - 1	CCB47	-175	-0.0748	-7471B	PPB		15 Jun 2018 16:46:59	U	SMPL
J3486-07 - 1	DOCUMENTATION-10	14388	3.3226	-7471B	PPB		15 Jun 2018 16:49:16	U	SMPL
J3504-01 - 1	SB-44COMP	807	0.1543	-7471B	PPB		15 Jun 2018 16:51:33	U	SMPL
J3504-02 - 1	SB-46COMP	910	0.1783	-7471B	PPB		15 Jun 2018 16:53:48	U	SMPL
J3504-03 - 1	SB-51COMP	1076	0.2170	-7471B	PPB		15 Jun 2018 16:56:05	U	SMPL
J3505-01 - 1	SB-56-COMP	195	0.0115	-7471B	PPB		15 Jun 2018 16:58:22	U	SMPL
J3505-02 - 1	SB-55-COMP	138	-0.0018	-7471B	PPB		15 Jun 2018 17:00:39	U	SMPL
PB110277BL - 1	PBS	238	0.0215	-7471B	PPB		15 Jun 2018 17:02:55	U	SMPL
PB110277BS - 1	LCSS	16442	3.8018	-7471B	PPB		15 Jun 2018 17:05:13	U	SMPL
J3518-01 - 1	EPE-106-7(3-8)	1010	0.2016	-7471B	PPB		15 Jun 2018 17:07:30	U	SMPL
J3518-01DUP - 1	EPE-106-7(3-8)DUP	936	0.1844	-7471B	PPB		15 Jun 2018 17:09:47	U	SMPL
CCV48 - 1	CCV48	21396	4.9575	-7471B	PPB		15 Jun 2018 17:12:03	U	SMPL
CCB48 - 1	CCB48	-99	-0.0571	-7471B	PPB		15 Jun 2018 17:14:19	U	SMPL
J3518-01MS - 1	EPE-106-7(3-8)MS	17434	4.0332	-7471B	PPB		15 Jun 2018 17:16:36	U	SMPL
J3518-01MSD - 1	EPE-106-7(3-8)MSD	17286	3.9987	-7471B	PPB		15 Jun 2018 17:18:52	U	SMPL
J3518-02 - 1	EPE-106-7(60-62)	1246	0.2567	-7471B	PPB		15 Jun 2018 17:21:08	U	SMPL
J3518-03 - 1	EPE-106-7(65-67)	898	0.1755	-7471B	PPB		15 Jun 2018 17:23:25	U	SMPL
J3518-04 - 1	EPE-106-7(70-72)	1783	0.3820	-7471B	PPB		15 Jun 2018 17:25:42	U	SMPL
J3486-01LX5 - 1		4063	0.9139	-7471B	PPB		15 Jun 2018 17:27:59	U	SMPL
J3486-01A - 1		38932	9.0486	-7471B	PPB		15 Jun 2018 17:30:16	U	SMPL
J3518-01LX5 - 1		25	-0.0282	-7471B	PPB		15 Jun 2018 17:32:34	U	SMPL
J3518-01A - 1		18547	4.2929	-7471B	PPB		15 Jun 2018 17:34:53	U	SMPL
J3486-04DLX10 - 1		40055	9.3106	-7471B	PPB		15 Jun 2018 17:37:09	U	SMPL
CCV49 - 1	CCV49	21240	4.9211	-7471B	PPB		15 Jun 2018 17:39:26	U	SMPL
CCB49 - 1	CCB49	-143	-0.0674	-7471B	PPB		15 Jun 2018 17:41:43	U	SMPL
J3486-03DLX2 - 1		26602	6.1721	-7471B	PPB		15 Jun 2018 17:44:00	U	SMPL
J3486-05DLX2 - 1		23278	5.3966	-7471B	PPB		15 Jun 2018 17:46:17	U	SMPL
J3486-06DLX5 - 1		16020	3.7034	-7471B	PPB		15 Jun 2018 17:48:34	U	SMPL
CCV50 - 1	CCV50	21150	4.9002	-7471B	PPB		15 Jun 2018 17:50:51	U	SMPL
CCB50 - 1	CCB50	-95	-0.0562	-7471B	PPB		15 Jun 2018 17:53:07	U	SMPL