

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/15/2018 15:34

Test: Total CN mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.924	0.0	0.095	
ICB1	-1.327	0.0	0.003	
CCV1	251.334	0.0	0.233	
CCB1	-1.499	0.0	0.003	
PB110302BL	-1.425	0.0	0.003	
PB110302BS	99.734	0.0	0.095	
MIDPB110302	247.558	0.0	0.229	
J3509-01	7.158	0.0	0.011	
J3509-02	38.125	0.0	0.039	
J3509-03	35.765	0.0	0.037	
J3509-04	-1.267	0.0	0.004	
J3509-05	0.813	0.0	0.005	
J3509-05DUP	0.515	0.0	0.005	
J3552-01	-1.299	0.0	0.003	
CCV2	250.181	0.0	0.232	
CCB2	-0.798	0.0	0.004	
J3552-02	-0.779	0.0	0.004	
CCV3	256.990	0.0	0.238	
CCB3	-1.838	0.0	0.003	

N 19
Mean 67.203
SD 102.6889
CV% 152.80

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : _____

6/15/2018 12:03


Test Total CN

Accepted 6/15/2018 12:03

Factor 1101
 Bias 0.005

Coeff. of det. 0.999727

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-2.1367	0.0000	
2	5.0PPBCN	0.008	3.6395	5.0000	
3	10PPBCN	0.012	8.5135	10.0000	
4	50PPBCN	0.050	49.8688	50.0000	
5	100PPBCN	0.098	103.0023	100.0000	
6	250PPBCN	0.237	255.3132	250.0000	
7	500PPBCN	0.456	496.7993	500.0000	

Aquakem v. 7.2AQ1

Results from time period:

Fri Jun 15 15:07:20 2018

Fri Jun 15 15:33:30 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-2.1367	µg/l	6/15/2018 12:02:59	
5.0PPBCN	A	Total CN	P	3.6395	µg/l	6/15/2018 12:03:00	
10PPBCN	A	Total CN	P	8.5135	µg/l	6/15/2018 12:03:01	
50PPBCN	A	Total CN	P	49.8688	µg/l	6/15/2018 12:03:02	
100PPBCN	A	Total CN	P	103.0023	µg/l	6/15/2018 12:03:03	
250PPBCN	A	Total CN	P	255.3132	µg/l	6/15/2018 12:03:04	
500PPBCN	A	Total CN	P	496.7993	µg/l	6/15/2018 12:03:05	
ICV1	S	Total CN	P	98.9238	µg/l	6/15/2018 15:07:20	
ICB1	S	Total CN	P	-1.3274	µg/l	6/15/2018 15:07:21	
CCV1	S	Total CN	P	251.3343	µg/l	6/15/2018 15:07:22	
CCB1	S	Total CN	P	-1.4993	µg/l	6/15/2018 15:07:23	
PB110302BL	S	Total CN	P	-1.4252	µg/l	6/15/2018 15:07:24	
PB110302BS	S	Total CN	P	99.7336	µg/l	6/15/2018 15:07:25	
MIDPB110302	S	Total CN	P	247.5584	µg/l	6/15/2018 15:07:26	
J3509-01	S	Total CN	P	7.1585	µg/l	6/15/2018 15:07:27	
J3509-02	S	Total CN	P	38.1252	µg/l	6/15/2018 15:07:28	
J3509-03	S	Total CN	P	35.7647	µg/l	6/15/2018 15:07:29	
J3509-04	S	Total CN	P	-1.2674	µg/l	6/15/2018 15:07:30	
J3509-05	S	Total CN	P	0.8133	µg/l	6/15/2018 15:13:31	
J3509-05DUP	S	Total CN	P	0.5154	µg/l	6/15/2018 15:13:32	
J3552-01	S	Total CN	P	-1.2993	µg/l	6/15/2018 15:13:33	
CCV2	S	Total CN	P	250.1807	µg/l	6/15/2018 15:13:34	
CCB2	S	Total CN	P	-0.7978	µg/l	6/15/2018 15:13:35	
J3552-02	S	Total CN	P	-0.7786	µg/l	6/15/2018 15:13:36	
CCV3	S	Total CN	P	256.9897	µg/l	6/15/2018 15:13:37	
CCB3	S	Total CN	P	-1.838	µg/l	6/15/2018 15:13:38	