



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96052

Slope : 98.4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/15/2018	12:10
2	CAL2	1	Water			21.6	7.00	06/15/2018	12:12
3	CAL3	1	Water			21.4	10.03	06/15/2018	12:14
4	ICV	1	Water			21.7	7.00	06/15/2018	12:16
5	CCV1	1	Water			21.5	2.00	06/15/2018	12:18
6	J3541-01	1	Water			21.2	4.26	06/15/2018	12:20
7	J3541-01DUP	1	Water			21.3	4.27	06/15/2018	12:22
8	J3541-02	1	Water			21.2	4.50	06/15/2018	12:24
9	J3541-03	1	Water			21.0	4.76	06/15/2018	12:26
10	CCV2	1	Water			21.5	12.00	06/15/2018	12:28

WORKLIST(Hardcopy Internal Chain)

LB96052

WorkList Name : pH061518

WorkList ID : 113426

Date : 6/15/2018 11:53:43 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/14/2018	Water	J3541-01	pH	Cool 4 deg C	STER13	D41	CELL-A-POOL-WATER	06/14/2018	SM 4500-PH B
06/14/2018	Water	J3541-02	pH	Cool 4 deg C	STER13	D41	CELL-B-POOL-WATER	06/14/2018	SM 4500-PH B
06/14/2018	Water	J3541-03	pH	Cool 4 deg C	STER13	D41	CELL-C-POOL-WATER	06/14/2018	SM 4500-PH B

Date/Time 06/15/18 12:00 PM

Received by: SP

Relinquished by: CP

Date/Time 06/15/18 12:35 PM

Received by: CP

Relinquished by: SP