



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96055

Slope : 98.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2271
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/15/2018	16:35
2	CAL2	1	Water			21.6	7.00	06/15/2018	16:37
3	CAL3	1	Water			21.4	10.03	06/15/2018	16:39
4	ICV	1	Water			21.6	7.00	06/15/2018	16:41
5	CCV1	1	Water			21.5	2.00	06/15/2018	16:43
6	J3532-01	1	Water			22.3	8.50	06/15/2018	16:45
7	J3532-01DUP	1	Water			22.4	8.51	06/15/2018	16:47
8	CCV2	1	Water			21.5	12.00	06/15/2018	16:49

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH061518-1

WorkList ID : 113437

Date : 6/15/2018 3:34:49 PM

LB 96055

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/15/2018	Water	J3532-01	pH	Cool 4 deg C	PSEG01	D41	OILY-WATER-DRUMS-1-2	06/15/2018	9040C

Date/Time 06/15/18 4:30 PM
 Received by: SP
 Relinquished by: CP

Date/Time 06/15/18 5:00 PM
 Received by: CP
 Relinquished by: SP