



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/18/2018

TEMP1 IN: 103 06/18/2018 07:20 TEMP OUT: 104 06/18/2018 08:35

TEMP2 IN: 104 06/18/2018 09:15 TEMP OUT: 104 06/18/2018 10:25

TEMP3 IN: 105 06/18/2018 12:00 TEMP OUT: 104 06/18/2018 13:40

TEMP4 IN: 105 06/18/2018 14:30 TEMP OUT: 105 06/18/2018 16:10

Run Number: LB96068

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96068BL	LB96068BL	1.3877	1.3877	100	1.3877	1.3877	1.3877	0.0000	0
2	LB96068BS	LB96068BS	1.4003	1.4003	100	1.4501	1.4501	1.4501	0.0498	498
3	J3522-02	TPTW-01	1.3961	1.3961	100	1.4176	1.4176	1.4176	0.0215	215
4	J3525-01	TPTW-04	1.3978	1.3978	10	1.5107	1.5107	1.5107	0.1129	11290
5	J3532-01	OILY-WATER-DRUMS-1-2	1.3940	1.3940	100	1.4574	1.4574	1.4574	0.0634	634
6	J3532-01DUP	OILY-WATER-DRUMS-1-2DUP	1.4007	1.4007	100	1.4655	1.4655	1.4655	0.0648	648
7	J3534-02	EFF-WW	1.4064	1.4064	300	1.4142	1.4142	1.4142	0.0078	26
8	J3546-01	061418-DBRC-F-D-B	1.4030	1.4030	300	1.4180	1.4180	1.4180	0.0150	50
9	J3546-02	061418-DBRC-F-D-S	1.4030	1.4030	200	1.4148	1.4148	1.4148	0.0118	59
10	J3546-03	061418-DBRC-F-U-B	1.4064	1.4064	200	1.4234	1.4234	1.4234	0.0170	85
11	J3546-04	061418-DBRC-F-U-S	1.4018	1.4018	300	1.4124	1.4124	1.4124	0.0106	35.3
12	J3548-01	061318-DBRI-E-D-B	1.4067	1.4067	200	1.4216	1.4216	1.4216	0.0149	74.5
13	J3548-02	061318-DBRI-E-D-M	1.3947	1.3947	300	1.4085	1.4085	1.4085	0.0138	46
14	J3548-03	061318-DBRI-E-D-S	1.4002	1.4002	400	1.4202	1.4202	1.4202	0.0200	50
15	J3548-04	061318-DBRI-E-U-B	1.4040	1.4040	300	1.4201	1.4201	1.4201	0.0161	53.7



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 06/18/2018TEMP1 IN: 103 06/18/2018 07:20 TEMP OUT: 104 06/18/2018 08:35TEMP2 IN: 104 06/18/2018 09:15 TEMP OUT: 104 06/18/2018 10:25TEMP3 IN: 105 06/18/2018 12:00 TEMP OUT: 104 06/18/2018 13:40TEMP4 IN: 105 06/18/2018 14:30 TEMP OUT: 105 06/18/2018 16:10Run Number: LB96068BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J3548-05	061318-DBRI-E-U-M	1.3955	1.3955	300	1.4096	1.4096	1.4096	0.0141	47
17	J3548-06	061318-DBRI-E-U-S	1.3992	1.3992	200	1.4130	1.4130	1.4130	0.0138	69
18	J3548-06DUP	061318-DBRI-E-U-SDUP	1.3907	1.3907	200	1.4050	1.4050	1.4050	0.0143	71.5
19	J3557-02	001M-WILLETTS-PT-BLVD (JUNE)	1.4084	1.4084	100	1.4142	1.4142	1.4142	0.0058	58
20	J3557-03	002M-35TH-AVE (JUNE)	1.4011	1.4011	100	1.4058	1.4058	1.4058	0.0047	47

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061819

WorkList ID : 113442

Date : 6/18/2018 7:26:02 AM

1396068

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Water	J3522-02	TSS	Cool 4 deg C	DVIR01	D22	TPTW-01	06/13/2018	SM2540 D
06/21/2018	Water	J3525-01	TSS	Cool 4 deg C	DVIR01	D11	TPTW-04	06/14/2018	SM2540 D
06/20/2018	Water	J3532-01	TSS	Cool 4 deg C	PSEG01	D41	OILY-WATER-DRUMS-1-2	06/15/2018	SM2540 D
06/22/2018	Water	J3534-02	TSS	Cool 4 deg C	ARDM01	D51	EFF-VWV	06/15/2018	SM2540 D
06/21/2018	Water	J3546-01	TSS	Cool 4 deg C	HDR103	D21	061418-DBRC-F-D-B	06/14/2018	SM2540 D
06/21/2018	Water	J3546-02	TSS	Cool 4 deg C	HDR103	D21	061418-DBRC-F-D-S	06/14/2018	SM2540 D
06/21/2018	Water	J3546-03	TSS	Cool 4 deg C	HDR103	D21	061418-DBRC-F-U-B	06/14/2018	SM2540 D
06/21/2018	Water	J3546-04	TSS	Cool 4 deg C	HDR103	D21	061418-DBRC-F-U-S	06/14/2018	SM2540 D
06/20/2018	Water	J3548-01	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-D-B	06/13/2018	SM2540 D
06/20/2018	Water	J3548-02	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-D-M	06/13/2018	SM2540 D
06/20/2018	Water	J3548-03	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-D-S	06/13/2018	SM2540 D
06/20/2018	Water	J3548-04	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-U-B	06/13/2018	SM2540 D
06/20/2018	Water	J3548-05	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-U-M	06/13/2018	SM2540 D
06/20/2018	Water	J3548-06	TSS	Cool 4 deg C	HDR103	D31	061318-DBRI-E-U-S	06/13/2018	SM2540 D
06/20/2018	Water	J3557-02	TSS	Cool 4 deg C	TULL01	D41	001M-WILLETTS-PT-BLVD(JUN	06/14/2018	SM2540 D
06/20/2018	Water	J3557-03	TSS	Cool 4 deg C	TULL01	D41	002M-35TH-AVE(JUNE)	06/14/2018	SM2540 D

Date/Time 06/18/18, 09:00
 Received by: AS
 Relinquished by: CP

Date/Time 06/18/18, 13:10
 Received by: CP
 Relinquished by: AS