

LB 96070


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37027



A= 0.0000e+000

B= 2.7401e-004

C= -1.1329e-001

Rho= 0.9999775

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	416	0.000	416				
0.2	0.200	0.201	0.001	1148	0.0 %	1148				
2.5	2.500	2.513	0.013	9585	0.0 %	9585				
5.0	5.000	5.002	0.002	18668	0.0 %	18668				
7.5	7.500	7.450	-0.050	27604	0.0 %	27604				
10.0	10.000	10.032	0.032	37027	0.0 %	37027				

LB96070 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jun 2018 09:34:26

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	416	-	0.0000	7470A	PPB	18 Jun 2018 09:50:27	S	Std
0.2 - 1	50.2	1148	-	0.2000	7470A	PPB	18 Jun 2018 09:52:35	S	Std
2.5 - 1	52.5	9585	-	2.5000	7470A	PPB	18 Jun 2018 09:54:43	S	Std
5.0 - 1	55	18668	-	5.0000	7470A	PPB	18 Jun 2018 09:56:51	S	Std
7.5 - 1	57.5	27604	-	7.5000	7470A	PPB	18 Jun 2018 09:58:59	S	Std
10.0 - 1	510	37027	-	10.0000	7470A	PPB	18 Jun 2018 10:03:43	S	Std
ICV06 - 1	ICV06	14379	3.8267	-	7470A	PPB	18 Jun 2018 10:06:26	U	SMPL
ICB06 - 1	ICB06	304	-0.0300	-	7470A	PPB	18 Jun 2018 10:08:32	U	SMPL
CCV51 - 1	CCV51	19248	5.1609	-	7470A	PPB	18 Jun 2018 10:10:39	U	SMPL
CCB51 - 1	CCB51	295	-0.0325	-	7470A	PPB	18 Jun 2018 10:12:46	U	SMPL
CRA - 1	CRA	1126	0.1952	-	7470A	PPB	18 Jun 2018 10:14:53	U	SMPL
HighStd - 1	HighStd	37396	10.1336	-	7470A	PPB	18 Jun 2018 10:17:00	U	SMPL
ChkStd - 1	ChkStd	25637	6.9115	-	7470A	PPB	18 Jun 2018 10:19:07	U	SMPL
PB110308BL - 1	PBW	222	-0.0525	-	7470A	PPB	18 Jun 2018 10:25:27	U	SMPL
PB110308BS - 1	LCSW	14251	3.7916	-	7470A	PPB	18 Jun 2018 10:27:34	U	SMPL
PB110308TB - 1	PB110254TB	86	-0.0897	-	7470A	PPB	18 Jun 2018 10:29:41	U	SMPL
J3464-07 - 1	SB-38-COMP	157	-0.0703	-	7470A	PPB	18 Jun 2018 10:31:49	U	SMPL
J3464-07DUP - 1	SB-38-COMPDUP	170	-0.0667	-	7470A	PPB	18 Jun 2018 10:33:57	U	SMPL
J3464-07MS - 1	SB-38-COMPMS	15091	4.0218	-	7470A	PPB	18 Jun 2018 10:36:06	U	SMPL
J3464-07MSD - 1	SB-38-COMPMSD	14777	3.9358	-	7470A	PPB	18 Jun 2018 10:38:13	U	SMPL
CCV52 - 1	CCV52	19235	5.1573	-	7470A	PPB	18 Jun 2018 10:40:20	U	SMPL
CCB52 - 1	CCB52	296	-0.0322	-	7470A	PPB	18 Jun 2018 10:42:26	U	SMPL
J3464-08 - 1	SB-39-COMP	141	-0.0747	-	7470A	PPB	18 Jun 2018 10:44:33	U	SMPL
J3464-09 - 1	SB-40-COMP	141	-0.0747	-	7470A	PPB	18 Jun 2018 10:46:40	U	SMPL
J3464-10 - 1	SB-41-COMP	143	-0.0741	-	7470A	PPB	18 Jun 2018 10:48:47	U	SMPL
J3464-11 - 1	SB-42-COMP	172	-0.0662	-	7470A	PPB	18 Jun 2018 10:50:54	U	SMPL
J3464-12 - 1	SB-43-COMP	165	-0.0681	-	7470A	PPB	18 Jun 2018 10:53:02	U	SMPL
J3498-12 - 1	TP-E-S	227	-0.0511	-	7470A	PPB	18 Jun 2018 10:55:09	U	SMPL
J3498-16 - 1	TP-D-S	239	-0.0478	-	7470A	PPB	18 Jun 2018 10:57:18	U	SMPL
J3501-04 - 1	TP-10-S	216	-0.0541	-	7470A	PPB	18 Jun 2018 10:59:26	U	SMPL
J3501-08 - 1	TP-10-D	179	-0.0642	-	7470A	PPB	18 Jun 2018 11:01:34	U	SMPL
J3501-12 - 1	TP-9	188	-0.0618	-	7470A	PPB	18 Jun 2018 11:03:41	U	SMPL
CCV53 - 1	CCV53	19209	5.1502	-	7470A	PPB	18 Jun 2018 11:05:48	U	SMPL
CCB53 - 1	CCB53	315	-0.0270	-	7470A	PPB	18 Jun 2018 11:07:54	U	SMPL
J3504-04 - 1	SB-44COMP	170	-0.0667	-	7470A	PPB	18 Jun 2018 11:10:01	U	SMPL
J3504-05 - 1	SB-46COMP	174	-0.0656	-	7470A	PPB	18 Jun 2018 11:12:08	U	SMPL
J3504-06 - 1	SB-51COMP	224	-0.0519	-	7470A	PPB	18 Jun 2018 11:14:14	U	SMPL
J3505-03 - 1	SB-56-COMP	252	-0.0442	-	7470A	PPB	18 Jun 2018 11:16:21	U	SMPL
J3505-04 - 1	SB-55-COMP	320	-0.0256	-	7470A	PPB	18 Jun 2018 11:18:29	U	SMPL
PB110309BL - 1	PBW	360	-0.0146	-	7470A	PPB	18 Jun 2018 11:20:36	U	SMPL
PB110309BS - 1	LCSW	13053	3.4634	-	7470A	PPB	18 Jun 2018 11:22:45	U	SMPL
PB110309TB - 1	PB110274TB	99	-0.0862	-	7470A	PPB	18 Jun 2018 11:24:53	U	SMPL
J3508-02 - 1	KGS-SIDE-1-A	202	-0.0579	-	7470A	PPB	18 Jun 2018 11:27:01	U	SMPL
J3508-02DUP - 1	KGS-SIDE-1-ADUP	257	-0.0429	-	7470A	PPB	18 Jun 2018 11:29:08	U	SMPL
CCV54 - 1	CCV54	19314	5.1789	-	7470A	PPB	18 Jun 2018 11:31:16	U	SMPL
CCB54 - 1	CCB54	323	-0.0248	-	7470A	PPB	18 Jun 2018 11:33:22	U	SMPL
J3508-02MS - 1	KGS-SIDE-1-AMS	15058	4.0128	-	7470A	PPB	18 Jun 2018 11:35:30	U	SMPL
J3508-02MSD - 1	KGS-SIDE-1-AMSD	14519	3.8651	-	7470A	PPB	18 Jun 2018 11:37:37	U	SMPL
J3508-05 - 1	KGS-SIDE-2-3-A	56	-0.0979	-	7470A	PPB	18 Jun 2018 11:39:45	U	SMPL
J3512-02 - 1	FADW4710L	216	-0.0541	-	7470A	PPB	18 Jun 2018 11:41:53	U	SMPL
PB110311BL - 1	PBW	198	-0.0590	-	7470A	PPB	18 Jun 2018 11:44:00	U	SMPL
PB110311BS - 1	LCSW	14092	3.7481	-	7470A	PPB	18 Jun 2018 11:46:08	U	SMPL
J3508-01 - 1	KGS-SIDE-1	150	-0.0722	-	7470A	PPB	18 Jun 2018 11:48:16	U	SMPL
J3508-04 - 1	KGS-SIDE-2-3	153	-0.0714	-	7470A	PPB	18 Jun 2018 11:50:25	U	SMPL
J3509-01 - 1	MWHR3-8-061318	193	-0.0604	-	7470A	PPB	18 Jun 2018 11:52:33	U	SMPL
J3509-02 - 1	J3509-05MS	13331	3.5395	-	7470A	PPB	18 Jun 2018 11:54:40	U	SMPL
CCV55 - 1	CCV55	19198	5.1472	-	7470A	PPB	18 Jun 2018 11:56:47	U	SMPL
CCB55 - 1	CCB55	302	-0.0305	-	7470A	PPB	18 Jun 2018 11:58:54	U	SMPL
J3509-03 - 1	J3509-05MSD	14363	3.8223	-	7470A	PPB	18 Jun 2018 12:01:02	U	SMPL
J3509-04 - 1	DUP-061318	117	-0.0812	-	7470A	PPB	18 Jun 2018 12:03:10	U	SMPL
J3509-05 - 1	MWHR2-6-061318	121	-0.0801	-	7470A	PPB	18 Jun 2018 12:05:18	U	SMPL
J3509-05DUP - 1	MWHR2-6-061318DUP	141	-0.0747	-	7470A	PPB	18 Jun 2018 12:07:26	U	SMPL
J3512-01 - 1	FADW4710L	941	0.1446	-	7470A	PPB	18 Jun 2018 12:09:34	U	SMPL
J3523-04 - 1	SOIL-FIELD-BLANK	10	-0.1105	-	7470A	PPB	18 Jun 2018 12:11:42	U	SMPL
J3528-03 - 1	TPTW-01	960	0.1498	-	7470A	PPB	18 Jun 2018 12:13:50	U	SMPL
J3528-05 - 1	GW-FIELD-BLANK	24	-0.1067	-	7470A	PPB	18 Jun 2018 12:15:58	U	SMPL
J3529-02 - 1	TPTW-03	1734	0.3618	-	7470A	PPB	18 Jun 2018 12:18:06	U	SMPL
J3529-04 - 1	TPTW-02	7232	1.8684	-	7470A	PPB	18 Jun 2018 12:20:14	U	SMPL
CCV56 - 1	CCV56	17966	4.8096	-	7470A	PPB	18 Jun 2018 12:22:22	U	SMPL
CCB56 - 1	CCB56	302	-0.0305	-	7470A	PPB	18 Jun 2018 12:24:30	U	SMPL
J3529-06 - 1	TPTW-04	39275	10.6485	-	7470A	PPB	18 Jun 2018 12:26:38	U	SMPL
J3529-09 - 1	TPTW-04MS	18575	4.9765	-	7470A	PPB	18 Jun 2018 12:28:46	U	SMPL
J3529-10 - 1	TPTW-04MSD	18829	5.0461	-	7470A	PPB	18 Jun 2018 12:30:54	U	SMPL
J3532-01 - 1	OILY-WATER-DRUMS-1-2	1703	0.3534	-	7470A	PPB	18 Jun 2018 12:33:02	U	SMPL
J3552-01 - 1	MWHR2-10-061418	89	-0.0889	-	7470A	PPB	18 Jun 2018 12:35:10	U	SMPL
J3552-02 - 1	MWHR2-8-061418	43	-0.1015	-	7470A	PPB	18 Jun 2018 12:37:18	U	SMPL
J3552-03 - 1	MWHR3-14-061418	43	-0.1015	-	7470A	PPB	18 Jun 2018 12:39:26	U	SMPL
J3552-04 - 1	MWHR2-9-061418	475	0.0169	-	7470A	PPB	18 Jun 2018 12:41:34	U	SMPL
CCV57 - 1	CCV57	18718	5.0156	-	7470A	PPB	18 Jun 2018 12:43:42	U	SMPL
CCB57 - 1	CCB57	313	-0.0275	-	7470A	PPB	18 Jun 2018 12:45:50	U	SMPL
J3464-07LX5 - 1		342	-0.0196	-	7470A	PPB	18 Jun 2018 12:47:58	U	SMPL

LB96070 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Jun 2018 09:34:26

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc.	Method	Units	Date	Type	Type
J3464-07A - 1		14108	3.7524		-7470A	PPB	18 Jun 2018 12:52:49	U	SMPL
J3508-02LX5 - 1		253	-0.0440		-7470A	PPB	18 Jun 2018 12:54:57	U	SMPL
J3508-02A - 1		13993	3.7209		-7470A	PPB	18 Jun 2018 12:57:04	U	SMPL
J3509-05LX5 - 1		241	-0.0473		-7470A	PPB	18 Jun 2018 12:59:12	U	SMPL
J3509-05A - 1		13870	3.6872		-7470A	PPB	18 Jun 2018 13:01:20	U	SMPL
CCV58 - 1	CCV58	17513	4.6855		-7470A	PPB	18 Jun 2018 13:03:27	U	SMPL
CCB58 - 1	CCB58	264	-0.0409		-7470A	PPB	18 Jun 2018 13:05:35	U	SMPL