



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB96078

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	6	27.0	80.6	80.9	06/18/2018	09:10
2	J3508-02		1	7	>100.0	>212.0	>212.0	06/18/2018	09:35
3	J3508-02DUP		1	7	>100.0	>212.0	>212.0	06/18/2018	09:55
4	J3508-05		1	6	>100.0	>212.0	>212.0	06/18/2018	10:15
5	J3512-02		1	7	>100.0	>212.0	>212.0	06/18/2018	10:45
6	J3532-02		1	8	>100.0	>212.0	>212.0	06/18/2018	11:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB96078

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP-061818

WorkList ID : 113462

Date : 6/18/2018 8:19:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2018	Water	J3508-02 C	Flash Point	Cool 4 deg C	PSEG01	D31	KGS-SIDE-1-A	06/14/2018	1010A
06/19/2018	Water	J3508-05 C	Flash Point	Cool 4 deg C	PSEG01	D31	KGS-SIDE-2-3-A	06/14/2018	1010A
06/19/2018	Water	J3512-02 B	Flash Point	Cool 4 deg C	PSEG01	D32	FADW4710L	06/14/2018	1010A
06/20/2018	Water	J3532-02 C	Flash Point	Cool 4 deg C	PSEG01	D41	OILY-WATER-DRUMS-1-2	06/15/2018	1010A

Date/Time 06/18/18 8:30 AM

Received by: KS

Relinquished by: CP

Date/Time 06/18/18 11:15 AM

Received by: CP

Relinquished by: KS