

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB96082

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP65303
calibration std. phosphate 0.5 ppm	WP65302
calibration std. phosphate 0.3 ppm	WP65301
calibration std. 0 ppm	WP65295
phosphate ICV std.	WP65308
phosphate CCV std.	WP65307
calibration std. phosphate 0.1 ppm	WP65300
calibration std. phosphate 0.05 ppm	WP65299
Combined reagent	WP65407
11 N sulfuric acid	WP61601
Phenolphthalein indicator	WP61777
Sodium hydroxide, 1N	WP61776

Intercept: -0.001Slope: 0.6502Regression: 0.999847

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	06/18/2018	13:29
2	CAL2	0.05	1	50	50	0.033	06/18/2018	13:29
3	CAL3	0.10	1	50	50	0.065	06/18/2018	13:30
4	CAL4	0.30	1	50	50	0.186	06/18/2018	13:30
5	CAL5	0.50	1	50	50	0.329	06/18/2018	13:31
6	CAL6	1.00	1	50	50	0.649	06/18/2018	13:31



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB96082

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.324	0.500	06/18/2018	13:31
2	ICB		1	50	50	0.000	0.002	06/18/2018	13:32
3	CCV1	0.50	1	50	50	0.327	0.504	06/18/2018	13:32
4	CCB1		1	50	50	0.000	0.002	06/18/2018	13:33
5	PB110360BL		1	50	50	0.000	0.002	06/18/2018	13:33
6	PB110360BS	0.50	1	50	50	0.296	0.457	06/18/2018	13:34
7	J3515-01		1	50	50	0.655	1.009	06/18/2018	13:34
8	J3515-01DUP		1	50	50	0.653	1.006	06/18/2018	13:35
9	J3515-01MS	0.50	1	50	50	0.712	1.097	06/18/2018	13:36
10	J3515-01MSD	0.50	1	50	50	0.709	1.092	06/18/2018	13:35
11	J3515-01		2	50	50	0.320	0.494	06/18/2018	13:46
12	J3515-01DUP		2	50	50	0.311	0.480	06/18/2018	13:46
13	J3515-01MS	0.50	2	50	50	0.498	0.767	06/18/2018	13:47
14	J3515-01MSD	0.50	2	50	50	0.502	0.774	06/18/2018	13:47
15	CCV2	0.50	1	50	50	0.328	0.506	06/18/2018	13:49
16	CCB2		1	50	50	0.000	0.002	06/18/2018	13:49