

7471B


LB96090

INSTRUMENT ID: CY1

Linear

 μ Abs.:

41183



A= 0.0000e+000

B= 2.4732e-004

C= -5.5204e-002

Rho= 0.9997054

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.047	0.047	413	0.000	413				
0.2	0.200	0.199	-0.001	1026	0.0 %	1026				
2.5	2.500	2.524	0.024	10429	0.0 %	10429				
5.0	5.000	4.965	-0.035	20297	0.0 %	20297				
7.5	7.500	7.335	-0.165	29882	0.0 %	29882				
10.0	10.000	10.130	0.130	41183	0.0 %	41183				

LB96090 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 18 Jun 2018 15:10:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	413	-	0.00007471B	PPB		18 Jun 2018 15:36:47	S	Std
0.2 - 1	50.2	1026	-	0.20007471B	PPB		18 Jun 2018 15:39:04	S	Std
2.5 - 1	52.5	10429	-	2.50007471B	PPB		18 Jun 2018 15:41:22	S	Std
5.0 - 1	55	20297	-	5.00007471B	PPB		18 Jun 2018 15:43:39	S	Std
7.5 - 1	57.5	29882	-	7.50007471B	PPB		18 Jun 2018 15:45:56	S	Std
10.0 - 1	510	41183	-	10.00007471B	PPB		18 Jun 2018 15:48:12	S	Std
ICV07 - 1	ICV07	16238	3.9608	-7471B	PPB		18 Jun 2018 15:51:07	U	SMPL
ICB07 - 1	ICB07	399	0.0435	-7471B	PPB		18 Jun 2018 15:53:23	U	SMPL
CCV59 - 1	CCV59	21860	5.3513	-7471B	PPB		18 Jun 2018 15:55:39	U	SMPL
CCB59 - 1	CCB59	388	0.0408	-7471B	PPB		18 Jun 2018 15:57:54	U	SMPL
CRA - 1	CRA	991	0.1899	-7471B	PPB		18 Jun 2018 16:00:10	U	SMPL
HighStd - 1	HighStd	41145	10.1209	-7471B	PPB		18 Jun 2018 16:02:26	U	SMPL
ChkStd - 1	ChkStd	26083	6.3957	-7471B	PPB		18 Jun 2018 16:04:42	U	SMPL
PB110315BL - 1	PBS	15	-0.0515	-7471B	PPB		18 Jun 2018 16:07:00	U	SMPL
PB110315BS - 1	LCSS	14574	3.5493	-7471B	PPB		18 Jun 2018 16:09:19	U	SMPL
J3523-01 - 1	AREA1(S)	8462	2.0376	-7471B	PPB		18 Jun 2018 16:11:36	U	SMPL
J3523-01DUP - 1	AREA1(S)DUP	8718	2.1010	-7471B	PPB		18 Jun 2018 16:13:53	U	SMPL
J3523-02 - 1	J3523-01MS	23651	5.7942	-7471B	PPB		18 Jun 2018 16:16:11	U	SMPL
J3523-03 - 1	J3523-01MSD	23414	5.7356	-7471B	PPB		18 Jun 2018 16:18:28	U	SMPL
J3523-05 - 1	AREA1(D)	71993	17.7504	-7471B	PPB		18 Jun 2018 16:20:45	U	SMPL
CCV60 - 1	CCV60	21560	5.2771	-7471B	PPB		18 Jun 2018 16:23:02	U	SMPL
CCB60 - 1	CCB60	340	0.0289	-7471B	PPB		18 Jun 2018 16:25:20	U	SMPL
J3523-06 - 1	AREA2(S)	4863	1.1475	-7471B	PPB		18 Jun 2018 16:27:37	U	SMPL
J3523-07 - 1	AREA2(D)	47065	11.5851	-7471B	PPB		18 Jun 2018 16:29:53	U	SMPL
J3524-01 - 1	AREA-3(S)	9415	2.2733	-7471B	PPB		18 Jun 2018 16:32:08	U	SMPL
J3524-03 - 1	AREA-3(D)	187855	46.4058	-7471B	PPB		18 Jun 2018 16:34:26	U	SMPL
J3538-01 - 1	SS-T-4-STONE	736	0.1268	-7471B	PPB		18 Jun 2018 16:41:54	U	SMPL
J3538-07 - 1	SS-T-5-STONE	3291	0.7587	-7471B	PPB		18 Jun 2018 16:44:11	U	SMPL
J3551-04 - 1	OILY-SOIL	2203	0.4896	-7471B	PPB		18 Jun 2018 16:46:28	U	SMPL
J3554-04 - 1	A-3(5.0)	953	0.1805	-7471B	PPB		18 Jun 2018 16:48:45	U	SMPL
J3554-05 - 1	A-4(2.0)	747	0.1295	-7471B	PPB		18 Jun 2018 16:51:03	U	SMPL
J3554-06 - 1	A-5(2.5)	2303	0.5144	-7471B	PPB		18 Jun 2018 16:53:19	U	SMPL
CCV61 - 1	CCV61	21823	5.3421	-7471B	PPB		18 Jun 2018 17:08:21	U	SMPL
CCB61 - 1	CCB61	386	0.0403	-7471B	PPB		18 Jun 2018 17:10:37	U	SMPL
J3554-07 - 1	A-6(3.0)	2380	0.5334	-7471B	PPB		18 Jun 2018 17:12:54	U	SMPL
J3554-08 - 1	A-7(3.0)	1769	0.3823	-7471B	PPB		18 Jun 2018 17:15:10	U	SMPL
J3554-09 - 1	A-8(3.0)	2506	0.5646	-7471B	PPB		18 Jun 2018 17:17:26	U	SMPL
PB110316BL - 1	PBS	170	-0.0132	-7471B	PPB		18 Jun 2018 17:19:43	U	SMPL
PB110316BS - 1	LCSS	15983	3.8978	-7471B	PPB		18 Jun 2018 17:21:59	U	SMPL
J3240-03 - 1	BU-01-061518	1929	0.4219	-7471B	PPB		18 Jun 2018 17:24:16	U	SMPL
J3244-01 - 1	HD-01-061518	1876	0.4088	-7471B	PPB		18 Jun 2018 17:26:33	U	SMPL
J3566-03 - 1	TP-8-S	1767	0.3818	-7471B	PPB		18 Jun 2018 17:28:51	U	SMPL
J3566-07 - 1	TP-8-D	1906	0.4162	-7471B	PPB		18 Jun 2018 17:31:09	U	SMPL
J3566-11 - 1	TP-7	11491	2.7868	-7471B	PPB		18 Jun 2018 17:33:25	U	SMPL
CCV62 - 1	CCV62	22158	5.4250	-7471B	PPB		18 Jun 2018 17:35:41	U	SMPL
CCB62 - 1	CCB62	208	-0.0038	-7471B	PPB		18 Jun 2018 17:37:57	U	SMPL
J3566-15 - 1	TP-6-S	3599	0.8349	-7471B	PPB		18 Jun 2018 17:40:14	U	SMPL
J3566-19 - 1	TP-6-D	3485	0.8067	-7471B	PPB		18 Jun 2018 17:42:30	U	SMPL
J3567-03 - 1	TP-5-S	4472	1.0508	-7471B	PPB		18 Jun 2018 17:44:46	U	SMPL
J3567-07 - 1	TP-5-D	364	0.0348	-7471B	PPB		18 Jun 2018 17:47:03	U	SMPL
J3567-11 - 1	TP-11-S	1497	0.3150	-7471B	PPB		18 Jun 2018 17:49:20	U	SMPL
J3567-11DUP - 1	TP-11-SDUP	1750	0.3776	-7471B	PPB		18 Jun 2018 17:51:37	U	SMPL
J3567-11MS - 1	TP-11-SMS	18112	4.4243	-7471B	PPB		18 Jun 2018 17:53:54	U	SMPL
J3567-11MSD - 1	TP-11-SMSD	17738	4.3318	-7471B	PPB		18 Jun 2018 17:56:11	U	SMPL
CCV63 - 1	CCV63	21832	5.3444	-7471B	PPB		18 Jun 2018 17:58:29	U	SMPL
CCB63 - 1	CCB63	199	-0.0060	-7471B	PPB		18 Jun 2018 18:00:45	U	SMPL
J3523-01LX5 - 1		2281	0.5089	-7471B	PPB		18 Jun 2018 18:03:02	U	SMPL
J3523-01A - 1		26334	6.4578	-7471B	PPB		18 Jun 2018 18:05:18	U	SMPL
J3567-11LX5 - 1		728	0.1248	-7471B	PPB		18 Jun 2018 18:07:35	U	SMPL
J3567-11A - 1		18799	4.5942	-7471B	PPB		18 Jun 2018 18:12:53	U	SMPL
J3523-05DLX5 - 1		16060	3.9168	-7471B	PPB		18 Jun 2018 18:15:09	U	SMPL
J3523-07DLX2 - 1		25799	6.3255	-7471B	PPB		18 Jun 2018 18:17:25	U	SMPL
J3524-03DLX10 - 1		20624	5.0456	-7471B	PPB		18 Jun 2018 18:19:42	U	SMPL
CCV64 - 1	CCV64	20932	5.1218	-7471B	PPB		18 Jun 2018 18:22:00	U	SMPL
CCB64 - 1	CCB64	284	0.0150	-7471B	PPB		18 Jun 2018 18:24:17	U	SMPL