



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96097

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/18/2018	16:20
2	CAL2	1	Water			21.8	7.00	06/18/2018	16:22
3	CAL3	1	Water			21.6	10.03	06/18/2018	16:24
4	ICV	1	Water			21.7	7.00	06/18/2018	16:26
5	CCV1	1	Water			21.6	2.00	06/18/2018	17:25
6	J3583-13	1	Water			23.7	6.99	06/18/2018	17:27
7	J3583-13DUP	1	Water			23.8	7.00	06/18/2018	17:29
8	J3583-14	1	Water			23.9	7.15	06/18/2018	17:31
9	J3583-15	1	Water			23.1	6.99	06/18/2018	17:33
10	J3583-16	1	Water			22.9	7.03	06/18/2018	17:35
11	J3583-17	1	Water			22.9	6.91	06/18/2018	17:37
12	J3583-18	1	Water			23.1	7.02	06/18/2018	17:39
13	J3583-19	1	Water			23.3	6.97	06/18/2018	17:41
14	J3583-20	1	Water			23.1	7.04	06/18/2018	17:43
15	J3583-21	1	Water			23.1	7.06	06/18/2018	17:45
16	CCV2	1	Water			21.8	12.00	06/18/2018	17:47
17	J3583-22	1	Water			23.4	7.11	06/18/2018	17:49
18	J3583-23	1	Water			23.4	7.07	06/18/2018	17:51
19	J3583-24	1	Water			23.4	7.09	06/18/2018	17:53
20	J3583-24DUP	1	Water			23.5	7.10	06/18/2018	17:55

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.9	2.00	06/18/2018	17:57

WORKLIST(Hardcopy Internal Chain)

UB96097

WorkList Name : pH061818

WorkList ID : 113500

Date : 6/18/2018 5:00:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2018	Water	J3583-13	pH	Cool 4 deg C	WARR01	D62	DW-1-MECH-RM-PT-SOURCI	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-14	pH	Cool 4 deg C	WARR01	D62	DW-1F-MECH-RM-30-FLUSH	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-15	pH	Cool 4 deg C	WARR01	D62	DW-2-1ST-FL-LOBBY-MENS-I	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-16	pH	Cool 4 deg C	WARR01	D62	DW-2F-1ST-FL-LOBBY-30-FL	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-17	pH	Cool 4 deg C	WARR01	D62	DW-3-2ND-FL-RM-214	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-18	pH	Cool 4 deg C	WARR01	D62	DW-3F-2ND-FL-RM-30-FLUSI	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-19	pH	Cool 4 deg C	WARR01	D62	DW-4-4TH-FL-RM-413	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-20	pH	Cool 4 deg C	WARR01	D62	DW-4F-4TH-FL-RM-30-FLUST	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-21	pH	Cool 4 deg C	WARR01	D62	DW-5-7TH-FL-RM-708	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-22	pH	Cool 4 deg C	WARR01	D62	DW-5F-7TH-FL-RM-30-FLUST	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-23	pH	Cool 4 deg C	WARR01	D62	DW-6-10TH-FL-RM-1002	06/18/2018	SM 4500-PH B
06/18/2018	Water	J3583-24	pH	Cool 4 deg C	WARR01	D62	DW-6F-10TH-FL-RM-30-FLUS	06/18/2018	SM 4500-PH B

Date/Time 06/18/18 5:05

Received by: SP

Relinquished by: CP

Date/Time 06/18/18 6:05 PM

Received by: CP

Relinquished by: SP