

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 06/20/2018

TEMP1 IN: 103 06/20/2018 07:10 TEMP OUT: 104 06/20/2018 08:20

TEMP2 IN: 104 06/20/2018 09:00 TEMP OUT: 104 06/20/2018 10:05

TEMP3 IN: 105 06/20/2018 11:30 TEMP OUT: 105 06/20/2018 13:10

TEMP4 IN: 105 06/20/2018 14:00 TEMP OUT: 105 06/20/2018 15:40

Run Number: LB96106

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB96106BL	LB96106BL	1.3985	1.3985	100	1.3985	1.3985	1.3985	0.0000	0
2	LB96106BS	LB96106BS	1.4126	1.4126	100	1.4661	1.4661	1.4661	0.0535	535
3	J3570-05	20180375-COMPOSITE-CS-1A-EL	1.4076	1.4076	70	1.4209	1.4209	1.4209	0.0133	190
4	J3570-06	20180376-COMPOSITE-CS-1B-EL	1.4012	1.4012	200	1.4176	1.4176	1.4176	0.0164	82
5	J3587-01	061818-DBRI-E-D-B	1.4018	1.4018	200	1.4200	1.4200	1.4200	0.0182	91
6	J3587-01DUP	061818-DBRI-E-D-BDUP	1.3999	1.3999	200	1.4180	1.4180	1.4180	0.0181	90.5
7	J3587-02	061818-DBRI-E-D-S	1.3983	1.3983	300	1.4159	1.4159	1.4159	0.0176	58.7
8	J3587-03	061818-DBRI-E-U-B	1.4029	1.4029	200	1.4185	1.4185	1.4185	0.0156	78
9	J3587-04	061818-DBRI-E-U-S	1.4030	1.4030	300	1.4135	1.4135	1.4135	0.0105	35
10	J3587-05	061818-TIPC-F-D-B	1.4032	1.4032	200	1.4162	1.4162	1.4162	0.0130	65
11	J3587-06	061818-TIPC-F-D-S	1.3996	1.3996	300	1.4137	1.4137	1.4137	0.0141	47
12	J3587-07	061818-TIPC-F-U-B	1.4032	1.4032	300	1.4170	1.4170	1.4170	0.0138	46
13	J3587-08	061818-TPIC-F-U-S	1.4027	1.4027	300	1.4159	1.4159	1.4159	0.0132	44



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 06/20/2018TEMP1 IN: 103 06/20/2018 07:10 TEMP OUT: 104 06/20/2018 08:20TEMP2 IN: 104 06/20/2018 09:00 TEMP OUT: 104 06/20/2018 10:05TEMP3 IN: 105 06/20/2018 11:30 TEMP OUT: 105 06/20/2018 13:10TEMP4 IN: 105 06/20/2018 14:00 TEMP OUT: 105 06/20/2018 15:40Run Number: LB96106BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
A = Sample Volume (ml)						Weight (g) = C - B				
B = Empty Dish Weight (g)						Result mg/L = $\frac{D}{A} \times 1000 \times 1000$				
C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C (g)										
D = Weight (g)										

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS061918

WorkList ID : 113502

Date : 6/19/2018 7:26:04 AM

2396106

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/22/2018	Water	J3570-05 E	TSS	Cool 4 deg C	AQUA04	D31	20180375-COMPOSITE-CS-1	06/15/2018	SM2540 D
06/22/2018	Water	J3570-06 E	TSS	Cool 4 deg C	AQUA04	D31	20180376-COMPOSITE-CS-1	06/15/2018	SM2540 D
06/25/2018	Water	J3587-01 C	TSS	Cool 4 deg C	HDR103	D32	061818-DBRI-E-D-B	06/18/2018	SM2540 D
06/25/2018	Water	J3587-02 C	TSS	Cool 4 deg C	HDR103	D32	061818-DBRI-E-D-S	06/18/2018	SM2540 D
06/25/2018	Water	J3587-03 C	TSS	Cool 4 deg C	HDR103	D32	061818-DBRI-E-U-B	06/18/2018	SM2540 D
06/25/2018	Water	J3587-04 C	TSS	Cool 4 deg C	HDR103	D32	061818-DBRI-E-U-S	06/18/2018	SM2540 D
06/25/2018	Water	J3587-05 C	TSS	Cool 4 deg C	HDR103	D32	061818-TIPC-F-D-B	06/18/2018	SM2540 D
06/25/2018	Water	J3587-06 C	TSS	Cool 4 deg C	HDR103	D32	061818-TIPC-F-D-S	06/18/2018	SM2540 D
06/25/2018	Water	J3587-07 C	TSS	Cool 4 deg C	HDR103	D32	061818-TIPC-F-U-B	06/18/2018	SM2540 D
06/25/2018	Water	J3587-08 C	TSS	Cool 4 deg C	HDR103	D32	061818-TIPC-F-U-S	06/18/2018	SM2540 D

Filter Paper ID:-9823888

Date/Time 06/20/18, 08:00
 Received by: AS
 Relinquished by: CS

Date/Time 06/20/18, 12:00
 Received by: CS
 Relinquished by: AS