

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : As

6/19/2018 10:24

Test: Phenolics-mg/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.061	0.0	0.111	
ICB1	0.007	0.0	0.004	
CCV1	1.006	0.0	0.105	
CCB1	0.003	0.0	0.004	
PB110365BL	-0.002	0.0	0.003	
PB110365BS	1.062	0.0	0.111	
J3503-01	0.001	0.0	0.004	
J3503-02	-0.004	0.0	0.003	
J3503-03	0.013	0.0	0.005	
J3509-01	0.003	0.0	0.004	
J3509-02MS	0.925	0.0	0.097	
J3509-03MSD	0.962	0.0	0.101	
J3509-05	0.003	0.0	0.004	
J3509-05DUP	0.003	0.0	0.004	
CCV2	0.981	0.0	0.103	
CCB2	0.000	0.0	0.004	
J3552-01	0.003	0.0	0.004	
J3552-05	0.005	0.0	0.004	
J3552-06	0.004	0.0	0.004	
CCV3	0.969	0.0	0.101	
CCB3	-0.000	0.0	0.004	

N	21
Mean	0.333
SD	0.4802
CV%	143.99

Aquakem v. 7.2AQ1

Results from time period:

Tue Jun 19 09:47:46 2018

Tue Jun 19 10:24:27 2018

AS

Sample Id	Sam/Ctr/c/	Test short nar	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.053	mg/l	6/19/2018 9:47:46	
0.1PPM	A	Phenolics-	P	0.0935	mg/l	6/19/2018 9:47:47	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	6/19/2018 9:47:48	
0.50PPM	A	Phenolics-	P	0.4987	mg/l	6/19/2018 9:47:49	
1.0PPM	A	Phenolics-	P	1.007	mg/l	6/19/2018 9:47:50	
2.0PPM	A	Phenolics-	P	1.997	mg/l	6/19/2018 9:47:51	
ICV1	S	Phenolics-	P	1.0612	mg/l	6/19/2018 10:16:35	
ICB1	S	Phenolics-	P	0.0067	mg/l	6/19/2018 10:16:36	
CCV1	S	Phenolics-	P	1.0061	mg/l	6/19/2018 10:16:37	
CCB1	S	Phenolics-	P	0.003	mg/l	6/19/2018 10:16:38	
PB110365BL	S	Phenolics-	P	-0.0024	mg/l	6/19/2018 10:16:39	
PB110365BS	S	Phenolics-	P	1.0616	mg/l	6/19/2018 10:16:40	
J3503-01	S	Phenolics-	P	0.0011	mg/l	6/19/2018 10:16:41	
J3503-02	S	Phenolics-	P	-0.0035	mg/l	6/19/2018 10:16:42	
J3503-03	S	Phenolics-	P	0.0128	mg/l	6/19/2018 10:16:43	
J3509-01	S	Phenolics-	P	0.0026	mg/l	6/19/2018 10:16:44	
J3509-02MS	S	Phenolics-	P	0.9249	mg/l	6/19/2018 10:16:45	
J3509-03MSD	S	Phenolics-	P	0.9618	mg/l	6/19/2018 10:16:46	
J3509-05	S	Phenolics-	P	0.0033	mg/l	6/19/2018 10:24:18	
J3509-05DUP	S	Phenolics-	P	0.003	mg/l	6/19/2018 10:24:19	
CCV2	S	Phenolics-	P	0.9809	mg/l	6/19/2018 10:24:20	
CCB2	S	Phenolics-	P	0.0002	mg/l	6/19/2018 10:24:21	
J3552-01	S	Phenolics-	P	0.0027	mg/l	6/19/2018 10:24:22	
J3552-05	S	Phenolics-	P	0.0046	mg/l	6/19/2018 10:24:23	
J3552-06	S	Phenolics-	P	0.0043	mg/l	6/19/2018 10:24:24	
CCV3	S	Phenolics-	P	0.9691	mg/l	6/19/2018 10:24:25	
CCB3	S	Phenolics-	P	-0.0004	mg/l	6/19/2018 10:24:26	

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

6/19/2018 9:49


Test Phenolics-

Accepted 6/19/2018 9:49

Factor 9.912
Bias 0.004

Coeff. of det. 0.999960

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0530	0.0500	
2	Phenol-2	0.013	0.0935	0.1000	
3	Phenol-2	0.029	0.2508	0.2500	
4	Phenol-2	0.054	0.4987	0.5000	
5	Phenol-2	0.105	1.0070	1.0000	
6	Phenol-2	0.205	1.9970	2.0000	