



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96110

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP65433
Turbidity std, 40NTU	WP65429
turbidity std 5 NTU	WP65432
turbidity std 20 NTU	WP65430
Turbidity std 1 NTU	WP65434
turbidity std 0 NTU	WP65435
turbidity std 10 NTU	WP65431
10 NTU Standard 500 ml	W2242

Intercept: 0.2352

Slope: 0.9621

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.368	06/19/2018	10:15
2	CAL2	1	1	1.220	06/19/2018	10:18
3	CAL3	5	1	4.880	06/19/2018	10:21
4	CAL4	10	1	9.960	06/19/2018	10:25
5	CAL5	20	1	19.300	06/19/2018	10:28
6	CAL6	40	1	38.800	06/19/2018	10:31



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB96110

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.360	9.484	06/19/2018	10:34
2	ICB		1	0.360	0.130	06/19/2018	10:37
3	CCV1	10	1	9.120	9.235	06/19/2018	10:40
4	CCB1		1	0.358	0.128	06/19/2018	10:43
5	LB96110BL		1	0.365	0.135	06/19/2018	10:46
6	J3583-13		1	0.470	0.244	06/19/2018	10:49
7	J3583-14		1	0.584	0.363	06/19/2018	10:52
8	J3583-15		1	0.452	0.225	06/19/2018	10:55
9	J3583-16		1	0.476	0.250	06/19/2018	10:58
10	J3583-17		1	0.463	0.237	06/19/2018	11:02
11	J3583-18		1	0.373	0.143	06/19/2018	11:05
12	J3583-19		1	0.440	0.213	06/19/2018	11:09
13	J3583-20		1	0.346	0.115	06/19/2018	11:12
14	J3583-21		1	0.310	0.078	06/19/2018	11:16
15	CCV2	10	1	9.420	9.547	06/19/2018	11:20
16	CCB2		1	0.369	0.139	06/19/2018	11:24
17	J3583-22		1	0.370	0.140	06/19/2018	11:27
18	J3583-23		1	0.355	0.125	06/19/2018	11:30
19	J3583-24		1	0.355	0.125	06/19/2018	11:34
20	J3583-24DUP		1	0.354	0.123	06/19/2018	11:39
21	CCV3	10	1	9.510	9.640	06/19/2018	11:43
22	CCB3		1	0.354	0.123	06/19/2018	11:46

WORKLIST(Hardcopy Internal Chain)

68 96110

WorkList Name : tur061918

WorkList ID : 113532

Date : 6/19/2018 9:23:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/20/2018	Water	J3583-13	Turbidity	Cool 4 deg C	WARR01	D62	DW-1-MECH-RM-PT-SOURCI	06/18/2018	SM2130 B
06/20/2018	Water	J3583-14	Turbidity	Cool 4 deg C	WARR01	D62	DW-1F-MECH-RM-30-FLUSH	06/18/2018	SM2130 B
06/20/2018	Water	J3583-15	Turbidity	Cool 4 deg C	WARR01	D62	DW-2-1ST-FL-I-LOBBY-MENS-I	06/18/2018	SM2130 B
06/20/2018	Water	J3583-16	Turbidity	Cool 4 deg C	WARR01	D62	DW-2F-1ST-FL-I-LOBBY-30-FL	06/18/2018	SM2130 B
06/20/2018	Water	J3583-17	Turbidity	Cool 4 deg C	WARR01	D62	DW-3-2ND-FL-RM-214	06/18/2018	SM2130 B
06/20/2018	Water	J3583-18	Turbidity	Cool 4 deg C	WARR01	D62	DW-3F-2ND-FL-RM-30-FLUSI	06/18/2018	SM2130 B
06/20/2018	Water	J3583-19	Turbidity	Cool 4 deg C	WARR01	D62	DW-4-4TH-FL-RM-413	06/18/2018	SM2130 B
06/20/2018	Water	J3583-20	Turbidity	Cool 4 deg C	WARR01	D62	DW-4F-4TH-FL-RM-30-FLUSI	06/18/2018	SM2130 B
06/20/2018	Water	J3583-21	Turbidity	Cool 4 deg C	WARR01	D62	DW-5-7TH-FL-RM-708	06/18/2018	SM2130 B
06/20/2018	Water	J3583-22	Turbidity	Cool 4 deg C	WARR01	D62	DW-5F-7TH-FL-RM-30-FLUSI	06/18/2018	SM2130 B
06/20/2018	Water	J3583-23	Turbidity	Cool 4 deg C	WARR01	D62	DW-6-10TH-FL-RM-1002	06/18/2018	SM2130 B
06/20/2018	Water	J3583-24	Turbidity	Cool 4 deg C	WARR01	D62	DW-6F-10TH-FL-RM-30-FLUS	06/18/2018	SM2130 B

Date/Time 6/19/18 9:46
 Received by: AV
 Relinquished by: AV

Date/Time 6-19-18 12:00
 Received by: AV
 Relinquished by: AV