

7470A


LB 96112

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40854



A= 0.0000e+000

B= 2.4729e-004

C= -3.5796e-002

Rho= 0.9996213

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.038	-0.038	-10	0.000	-10				
0.2	0.200	0.174	-0.026	849	0.0 %	849				
2.5	2.500	2.691	0.191	11026	0.0 %	11026				
5.0	5.000	4.900	-0.100	19961	0.0 %	19961				
7.5	7.500	7.406	-0.094	30092	0.0 %	30092				
10.0	10.000	10.067	0.067	40854	0.0 %	40854				

LB96112 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jun 2018 13:23:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-10	-	0.0000	7470A	PPB	19 Jun 2018 13:24:33	S	Std
0.2 - 1	50.2	849	-	0.2000	7470A	PPB	19 Jun 2018 13:26:41	S	Std
2.5 - 1	52.5	11026	-	2.5000	7470A	PPB	19 Jun 2018 13:28:50	S	Std
5.0 - 1	55	19961	-	5.0000	7470A	PPB	19 Jun 2018 13:30:58	S	Std
7.5 - 1	57.5	30092	-	7.5000	7470A	PPB	19 Jun 2018 13:35:47	S	Std
10.0 - 1	511	40854	-	10.0000	7470A	PPB	19 Jun 2018 13:42:51	S	Std
ICV09 - 1	ICV09	15220	3.7280	-7470A	PPB	PPB	19 Jun 2018 13:51:53	U	SMPL
ICB09 - 1	ICB09	-139	-0.0702	-7470A	PPB	PPB	19 Jun 2018 13:53:59	U	SMPL
CCV69 - 1	CCV69	20208	4.9615	-7470A	PPB	PPB	19 Jun 2018 13:56:06	U	SMPL
CCB69 - 1	CCB69	-151	-0.0731	-7470A	PPB	PPB	19 Jun 2018 13:58:12	U	SMPL
CRA - 1	CRA	826	0.1685	-7470A	PPB	PPB	19 Jun 2018 14:00:19	U	SMPL
HighStd - 1	HighStd	41062	10.1185	-7470A	PPB	PPB	19 Jun 2018 14:02:26	U	SMPL
ChkStd - 1	ChkStd	26262	6.4586	-7470A	PPB	PPB	19 Jun 2018 14:04:33	U	SMPL
PB110357BL - 1	PBW	-47	-0.0474	-7470A	PPB	PPB	19 Jun 2018 14:09:07	U	SMPL
PB110357BS - 1	LCSW	15080	3.6934	-7470A	PPB	PPB	19 Jun 2018 14:11:13	U	SMPL
PB110357TB - 1	PB110314TB	-130	-0.0679	-7470A	PPB	PPB	19 Jun 2018 14:13:21	U	SMPL
J3240-04 - 1	BU-01-061518	-51	-0.0484	-7470A	PPB	PPB	19 Jun 2018 14:15:29	U	SMPL
J3240-04DUP - 1	BU-01-061518DUP	-42	-0.0462	-7470A	PPB	PPB	19 Jun 2018 14:17:37	U	SMPL
J3240-04MS - 1	BU-01-061518MS	15612	3.8249	-7470A	PPB	PPB	19 Jun 2018 14:22:21	U	SMPL
J3240-04MSD - 1	BU-01-061518MSD	15445	3.7836	-7470A	PPB	PPB	19 Jun 2018 14:24:28	U	SMPL
CCV70 - 1	CCV70	19956	4.8992	-7470A	PPB	PPB	19 Jun 2018 14:26:35	U	SMPL
CCB70 - 1	CCB70	-169	-0.0776	-7470A	PPB	PPB	19 Jun 2018 14:28:41	U	SMPL
J3244-02 - 1	HD-01-061518	-51	-0.0484	-7470A	PPB	PPB	19 Jun 2018 14:30:49	U	SMPL
J3523-08 - 1	AREA1(S)	-58	-0.0501	-7470A	PPB	PPB	19 Jun 2018 14:32:55	U	SMPL
J3523-11 - 1	AREA1(D)	3	-0.0351	-7470A	PPB	PPB	19 Jun 2018 14:35:02	U	SMPL
J3523-13 - 1	AREA2(S)	-52	-0.0487	-7470A	PPB	PPB	19 Jun 2018 14:37:09	U	SMPL
J3523-14 - 1	AREA2(D)	-42	-0.0462	-7470A	PPB	PPB	19 Jun 2018 14:39:17	U	SMPL
J3524-02 - 1	AREA-3(S)	-41	-0.0459	-7470A	PPB	PPB	19 Jun 2018 14:41:24	U	SMPL
J3524-04 - 1	AREA-3(D)	-41	-0.0459	-7470A	PPB	PPB	19 Jun 2018 14:43:32	U	SMPL
J3538-06 - 1	SS-T-4-STONE	-39	-0.0454	-7470A	PPB	PPB	19 Jun 2018 14:45:40	U	SMPL
J3538-12 - 1	SS-T-5-STONE	-30	-0.0432	-7470A	PPB	PPB	19 Jun 2018 14:47:49	U	SMPL
J3540-01 - 1	SOMERSET-DRUM-1	-22	-0.0412	-7470A	PPB	PPB	19 Jun 2018 14:49:56	U	SMPL
CCV71 - 1	CCV71	20000	4.9101	-7470A	PPB	PPB	19 Jun 2018 14:52:03	U	SMPL
CCB71 - 1	CCB71	-166	-0.0768	-7470A	PPB	PPB	19 Jun 2018 14:54:10	U	SMPL
J3551-05 - 1	OILY-SOIL	-27	-0.0425	-7470A	PPB	PPB	19 Jun 2018 14:56:18	U	SMPL
J3566-04 - 1	TP-8-S	-41	-0.0459	-7470A	PPB	PPB	19 Jun 2018 14:58:25	U	SMPL
J3566-08 - 1	TP-8-D	-27	-0.0425	-7470A	PPB	PPB	19 Jun 2018 15:00:32	U	SMPL
J3566-12 - 1	TP-7	-33	-0.0440	-7470A	PPB	PPB	19 Jun 2018 15:02:39	U	SMPL
J3566-16 - 1	TP-6-S	-34	-0.0442	-7470A	PPB	PPB	19 Jun 2018 15:04:46	U	SMPL
J3566-20 - 1	TP-6-D	-36	-0.0447	-7470A	PPB	PPB	19 Jun 2018 15:06:54	U	SMPL
J3567-04 - 1	TP-5-S	-27	-0.0425	-7470A	PPB	PPB	19 Jun 2018 15:09:02	U	SMPL
J3567-08 - 1	TP-5-D	-11	-0.0385	-7470A	PPB	PPB	19 Jun 2018 15:11:10	U	SMPL
J3567-12 - 1	TP-11-S	-10	-0.0383	-7470A	PPB	PPB	19 Jun 2018 15:13:19	U	SMPL
CCV72 - 1	CCV72	19958	4.8997	-7470A	PPB	PPB	19 Jun 2018 15:15:26	U	SMPL
CCB72 - 1	CCB72	260	0.0285	-7470A	PPB	PPB	19 Jun 2018 15:17:33	U	SMPL
J3240-04LX5 - 1		-42	-0.0462	-7470A	PPB	PPB	19 Jun 2018 15:19:41	U	SMPL
J3240-04A - 1		14992	3.6716	-7470A	PPB	PPB	19 Jun 2018 15:21:48	U	SMPL
CCV73 - 1	CCV73	19846	4.8720	-7470A	PPB	PPB	19 Jun 2018 15:23:55	U	SMPL
CCB73 - 1	CCB73	-166	-0.0768	-7470A	PPB	PPB	19 Jun 2018 15:26:02	U	SMPL