



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96113

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.5	4.01	06/19/2018	09:20
2	CAL2	1	Water			21.4	7.00	06/19/2018	09:22
3	CAL3	1	Water			21.6	10.03	06/19/2018	09:24
4	ICV	1	Water			21.5	7.00	06/19/2018	09:26
5	CCV1	1	Water			21.4	2.00	06/19/2018	09:28
6	J3576-01	1	Solid	20.02	20	21.3	6.52	06/19/2018	09:30
7	J3576-01DUP	1	Solid	20.03	20	21.4	6.53	06/19/2018	09:32
8	J3584-04	1	Solid	20.02	20	21.7	7.56	06/19/2018	09:34
9	J3584-08	1	Solid	20.04	20	21.6	8.02	06/19/2018	09:36
10	J3584-12	1	Solid	20.02	20	21.5	6.75	06/19/2018	09:38
11	J3584-16	1	Solid	20.04	20	21.7	9.21	06/19/2018	09:40
12	J3584-20	1	Solid	20.02	20	22.9	7.62	06/19/2018	09:42
13	J3584-24	1	Solid	20.03	20	23.2	5.12	06/19/2018	09:44
14	CCV2	1	Water			21.6	12.00	06/19/2018	09:46

WORKLIST(Hardcopy Internal Chain)

1396113

WorkList Name : corrosivity061918

WorkList ID : 113530

Date : 6/19/2018 9:08:58 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2018	Solid	J3576-01	Corrosivity	Cool 4 deg C	PSEG01	D52	N-1366-OILY-DEBRIS	06/18/2018	9045C
06/18/2018	Solid	J3584-04	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-12-S	06/18/2018	9045C
06/18/2018	Solid	J3584-08	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-12-D	06/18/2018	9045C
06/18/2018	Solid	J3584-12	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-3	06/18/2018	9045C
06/18/2018	Solid	J3584-16	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-2-S	06/18/2018	9045C
06/18/2018	Solid	J3584-20	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-1-S	06/18/2018	9045C
06/18/2018	Solid	J3584-24	Corrosivity	Cool 4 deg C	PSEG01	D52	TP-1-D	06/18/2018	9045C

Date/Time 06/19/18 9:10 AM

Received by: SP

Relinquished by: CP

Date/Time 06/19/18 10:00 AM

Received by: CP

Relinquished by: SP