

7471B


LB 96122

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51089



A= 0.0000e+000

B= 1.9734e-004

C= -1.7700e-002

Rho= 0.9997905

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.012	-0.012	29	0.000	29				
0.2	0.200	0.192	-0.008	1062	0.0 %	1062				
2.5	2.500	2.622	0.122	13377	0.0 %	13377				
5.0	5.000	4.893	-0.107	24886	0.0 %	24886				
7.5	7.500	7.440	-0.060	37791	0.0 %	37791				
10.0	10.000	10.064	0.064	51089	0.0 %	51089				

LB96122 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jun 2018 15:43:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B	PPB		19 Jun 2018 15:46:16	S	Std
0.2 - 1	50.2	1062	-	0.20007471B	PPB		19 Jun 2018 15:48:33	S	Std
2.5 - 1	52.5	13377	-	2.50007471B	PPB		19 Jun 2018 15:50:51	S	Std
5.0 - 1	55	24886	-	5.00007471B	PPB		19 Jun 2018 15:56:19	S	Std
7.5 - 1	57.5	37791	-	7.50007471B	PPB		19 Jun 2018 16:03:51	S	Std
10.0 - 1	516	51089	-	10.00007471B	PPB		19 Jun 2018 16:08:49	S	Std
ICV10 - 1	ICV10	19970	3.9233	-7471B	PPB		19 Jun 2018 16:19:33	U	SMPL
ICB10 - 1	ICB10	-163	-0.0499	-7471B	PPB		19 Jun 2018 16:21:48	U	SMPL
CCV74 - 1	CCV74	26295	5.1715	-7471B	PPB		19 Jun 2018 16:24:04	U	SMPL
CCB74 - 1	CCB74	-188	-0.0548	-7471B	PPB		19 Jun 2018 16:26:20	U	SMPL
CRA - 1	CRA	1011	0.1818	-7471B	PPB		19 Jun 2018 16:28:37	U	SMPL
HighStd - 1	HighStd	51303	10.1066	-7471B	PPB		19 Jun 2018 16:30:53	U	SMPL
ChkStd - 1	ChkStd	33710	6.6348	-7471B	PPB		19 Jun 2018 16:37:27	U	SMPL
PB110371BL - 1	PBS	-245	-0.0660	-7471B	PPB		19 Jun 2018 16:39:43	U	SMPL
PB110371BS - 1	LCSS	17563	3.4483	-7471B	PPB		19 Jun 2018 16:49:07	U	SMPL
J3249-01 - 1	PL-01-061818-A	1847	0.3468	-7471B	PPB		19 Jun 2018 16:51:24	U	SMPL
J3249-01DUP - 1	PL-01-061818-ADUP	1601	0.2982	-7471B	PPB		19 Jun 2018 16:53:41	U	SMPL
J3249-01MS - 1	PL-01-061818-AMS	18421	3.6176	-7471B	PPB		19 Jun 2018 16:55:59	U	SMPL
J3249-01MSD - 1	PL-01-061818-AMSD	19525	3.8354	-7471B	PPB		19 Jun 2018 16:58:17	U	SMPL
J3249-03 - 1	PL-01-061818-B	1029	0.1854	-7471B	PPB		19 Jun 2018 17:00:32	U	SMPL
CCV75 - 1	CCV75	24467	4.8107	-7471B	PPB		19 Jun 2018 17:02:48	U	SMPL
CCB75 - 1	CCB75	-152	-0.0477	-7471B	PPB		19 Jun 2018 17:05:03	U	SMPL
J3249-05 - 1	PL-01-061818-C	3003	0.5749	-7471B	PPB		19 Jun 2018 17:07:20	U	SMPL
J3249-07 - 1	PL-01-061818-D	1922	0.3616	-7471B	PPB		19 Jun 2018 17:09:36	U	SMPL
J3249-09 - 1	PL-01-061818-E	5622	1.0918	-7471B	PPB		19 Jun 2018 17:11:52	U	SMPL
J3498-03 - 1	TP-G-S	755	0.1313	-7471B	PPB		19 Jun 2018 17:14:08	U	SMPL
J3498-07 - 1	TP-G-D	831	0.1463	-7471B	PPB		19 Jun 2018 17:16:25	U	SMPL
J3570-01 - 1	20180371-COMPOSITE-CS-1A	73836	14.5534	-7471B	PPB		19 Jun 2018 17:18:42	U	SMPL
J3570-02 - 1	20180372-COMPOSITE-CS-1B	34647	6.8197	-7471B	PPB		19 Jun 2018 17:21:00	U	SMPL
J3570-03 - 1	20180373-AMENDED-COMPOSITE-CS-1A	94845	18.6994	-7471B	PPB		19 Jun 2018 17:23:20	U	SMPL
J3570-04 - 1	20180374-AMENDED-COMPOSITE-CS-1B	27307	5.3712	-7471B	PPB		19 Jun 2018 17:25:38	U	SMPL
J3575-02 - 1	TP-9	1854	0.3482	-7471B	PPB		19 Jun 2018 17:27:57	U	SMPL
CCV76 - 1	CCV76	24942	4.9045	-7471B	PPB		19 Jun 2018 17:30:15	U	SMPL
CCB76 - 1	CCB76	-174	-0.0520	-7471B	PPB		19 Jun 2018 17:32:31	U	SMPL
J3575-06 - 1	TP-10	12026	2.3556	-7471B	PPB		19 Jun 2018 17:34:47	U	SMPL
J3575-10 - 1	TP-12	12126	2.3753	-7471B	PPB		19 Jun 2018 17:37:04	U	SMPL
J3575-14 - 1	TP-11-S	1537	0.2856	-7471B	PPB		19 Jun 2018 17:39:20	U	SMPL
J3575-18 - 1	TP-11-D	1882	0.3537	-7471B	PPB		19 Jun 2018 17:41:37	U	SMPL
J3575-21 - 1	TP-13	6022	1.1707	-7471B	PPB		19 Jun 2018 17:43:53	U	SMPL
PB110372BL - 1	PBS	-36	-0.0248	-7471B	PPB		19 Jun 2018 17:46:10	U	SMPL
PB110372BS - 1	LCSS	16494	3.2373	-7471B	PPB		19 Jun 2018 17:48:27	U	SMPL
J3573-02 - 1	SB-47-COMP	2715	0.5181	-7471B	PPB		19 Jun 2018 17:50:45	U	SMPL
J3573-03 - 1	SB-48-COMP	436	0.0683	-7471B	PPB		19 Jun 2018 17:53:02	U	SMPL
J3573-04 - 1	SB-49-COMP	398	0.0608	-7471B	PPB		19 Jun 2018 17:55:18	U	SMPL
CCV77 - 1	CCV77	25566	5.0276	-7471B	PPB		19 Jun 2018 17:57:35	U	SMPL
CCB77 - 1	CCB77	-149	-0.0471	-7471B	PPB		19 Jun 2018 17:59:51	U	SMPL
J3573-05 - 1	SB-50-COMP	385	0.0583	-7471B	PPB		19 Jun 2018 18:02:08	U	SMPL
J3573-06 - 1	SB-54-COMP	206	0.0230	-7471B	PPB		19 Jun 2018 18:04:24	U	SMPL
J3573-07 - 1	SB-45-COMP	1828	0.3430	-7471B	PPB		19 Jun 2018 18:06:40	U	SMPL
J3573-18 - 1	SB-52-COMP	442	0.0695	-7471B	PPB		19 Jun 2018 18:08:57	U	SMPL
J3573-19 - 1	SB-53-COMP	123	0.0066	-7471B	PPB		19 Jun 2018 18:11:14	U	SMPL
J3573-19DUP - 1	SB-53-COMPDUP	156	0.0131	-7471B	PPB		19 Jun 2018 18:13:31	U	SMPL
J3573-19MS - 1	SB-53-COMPMS	17974	3.5294	-7471B	PPB		19 Jun 2018 18:18:34	U	SMPL
J3573-19MSD - 1	SB-53-COMPMSD	17745	3.4842	-7471B	PPB		19 Jun 2018 18:20:52	U	SMPL
J3580-01 - 1	S-1	1698	0.3174	-7471B	PPB		19 Jun 2018 18:23:10	U	SMPL
J3580-02 - 1	S-2	1857	0.3488	-7471B	PPB		19 Jun 2018 18:25:26	U	SMPL
CCV78 - 1	CCV78	25653	5.0448	-7471B	PPB		19 Jun 2018 18:27:42	U	SMPL
CCB78 - 1	CCB78	-167	-0.0507	-7471B	PPB		19 Jun 2018 18:29:59	U	SMPL
J3580-03 - 1	S-3	1839	0.3452	-7471B	PPB		19 Jun 2018 18:32:15	U	SMPL
J3581-01 - 1	SSP-A-COMPOSITE	9019	1.7621	-7471B	PPB		19 Jun 2018 18:34:31	U	SMPL
J3581-04 - 1	SSP-B-COMPOSITE	5799	1.1267	-7471B	PPB		19 Jun 2018 18:36:47	U	SMPL
J3581-07 - 1	SSP-C-COMPOSITE	5976	1.1616	-7471B	PPB		19 Jun 2018 18:39:04	U	SMPL
J3584-03 - 1	TP-12-S	483	0.0776	-7471B	PPB		19 Jun 2018 18:41:20	U	SMPL
J3584-07 - 1	TP-12-D	3428	0.6588	-7471B	PPB		19 Jun 2018 18:43:37	U	SMPL
J3584-11 - 1	TP-3	604	0.1015	-7471B	PPB		19 Jun 2018 18:45:55	U	SMPL
J3584-15 - 1	TP-2-S	11137	2.1801	-7471B	PPB		19 Jun 2018 18:48:12	U	SMPL
J3584-19 - 1	TP-1-S	2929	0.5603	-7471B	PPB		19 Jun 2018 18:50:30	U	SMPL
J3584-23 - 1	TP-1-D	873	0.1546	-7471B	PPB		19 Jun 2018 18:52:47	U	SMPL
CCV79 - 1	CCV79	25722	5.0584	-7471B	PPB		19 Jun 2018 18:55:04	U	SMPL
CCB79 - 1	CCB79	-146	-0.0465	-7471B	PPB		19 Jun 2018 18:57:20	U	SMPL
J3249-01LX5 - 1		220	0.0257	-7471B	PPB		19 Jun 2018 18:59:37	U	SMPL
J3249-01A - 1		18999	3.7316	-7471B	PPB		19 Jun 2018 19:01:54	U	SMPL
J3573-19LX5 - 1		-111	-0.0396	-7471B	PPB		19 Jun 2018 19:04:11	U	SMPL
J3573-19A - 1		17379	3.4119	-7471B	PPB		19 Jun 2018 19:06:28	U	SMPL
J3570-01DLX5 - 1		15356	3.0127	-7471B	PPB		19 Jun 2018 19:12:01	U	SMPL
J3570-03DLX5 - 1		18245	3.5828	-7471B	PPB		19 Jun 2018 19:17:00	U	SMPL
CCV80 - 1	CCV80	25042	4.9242	-7471B	PPB		19 Jun 2018 19:19:16	U	SMPL
CCB80 - 1	CCB80	-151	-0.0475	-7471B	PPB		19 Jun 2018 19:21:33	U	SMPL