



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96128

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.6	4.01	06/19/2018	17:25
2	CAL2	1	Water			21.6	7.00	06/19/2018	17:27
3	CAL3	1	Water			21.8	10.03	06/19/2018	17:29
4	ICV	1	Water			21.7	7.00	06/19/2018	17:31
5	CCV1	1	Water			21.8	2.00	06/19/2018	18:01
6	J3589-01	1	Solid	20.02	20	23.9	8.61	06/19/2018	18:03
7	J3589-01DUP	1	Solid	20.03	20	24.0	8.62	06/19/2018	18:05
8	J3589-02	1	Solid	20.02	20	23.5	7.16	06/19/2018	18:07
9	J3589-03	1	Solid	20.03	20	23.2	8.21	06/19/2018	18:09
10	J3606-01	1	Solid	20.02	20	23.8	5.68	06/19/2018	18:11
11	J3611-05	1	Solid	20.04	20	23.1	6.72	06/19/2018	18:13
12	CCV2	1	Water			21.6	12.00	06/19/2018	18:15

WORKLIST(Hardcopy Internal Chain)

LB 96128

WorkList Name : pH061918

WorkList ID : 113551

Date : 6/19/2018 3:54:31 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2018	Solid	J3589-01	pH	Cool 4 deg C	EART13	D42	B-1	06/18/2018	9045C
06/18/2018	Solid	J3589-02	pH	Cool 4 deg C	EART13	D42	B-2	06/18/2018	9045C
06/18/2018	Solid	J3589-03	pH	Cool 4 deg C	EART13	D42	B-3	06/18/2018	9045C
06/19/2018	Solid	J3606-01	pH	Cool 4 deg C	TULL02	D41	FK-PS-01-030	06/19/2018	9045C
06/19/2018	Solid	J3611-05	pH	Cool 4 deg C	PSEG01	D62	N-1366-WATER	06/19/2018	9045C

Date/Time 06/19/18 5:10 PM
 Received by: SP
 Relinquished by: CP

Date/Time 06/19/18 6:25 PM
 Received by: CP
 Relinquished by: SP