

7470A


LB96138

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40508



A= 0.0000e+000

B= 2.4880e-004

C= -1.8653e-003

Rho= 0.9998255

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	40	0.000	40				
0.2	0.200	0.186	-0.014	756	0.0 %	756				
2.5	2.500	2.594	0.094	10434	0.0 %	10434				
5.0	5.000	4.916	-0.084	19765	0.0 %	19765				
7.5	7.500	7.420	-0.080	29830	0.0 %	29830				
10.0	10.000	10.076	0.076	40508	0.0 %	40508				

LB96138 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jun 2018 09:30:54

Sample ID	Extended ID	µ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0007470A	PPB	20 Jun 2018 09:34:58	S	Std
0.2 - 1	50.2	756	-	0.2007470A	PPB	20 Jun 2018 09:37:06	S	Std
2.5 - 1	52.5	10434	-	2.5007470A	PPB	20 Jun 2018 09:39:15	S	Std
5.0 - 1	55	19765	-	5.0007470A	PPB	20 Jun 2018 09:44:08	S	Std
7.5 - 1	57.5	29830	-	7.5007470A	PPB	20 Jun 2018 09:48:40	S	Std
10.0 - 1	59	40508	-	10.0007470A	PPB	20 Jun 2018 09:53:14	S	Std
ICV11 - 1	ICV11	15549	3.8667	-7470A	PPB	20 Jun 2018 10:00:29	U	SMPL
ICB11 - 1	ICB11	-164	-0.0427	-7470A	PPB	20 Jun 2018 10:02:35	U	SMPL
CCV81 - 1	CCV81	21571	5.3649	-7470A	PPB	20 Jun 2018 10:04:42	U	SMPL
CCB81 - 1	CCB81	-195	-0.0504	-7470A	PPB	20 Jun 2018 10:06:48	U	SMPL
CRA - 1	CRA	757	0.1865	-7470A	PPB	20 Jun 2018 10:08:55	U	SMPL
HighStd - 1	HighStd	39803	9.9009	-7470A	PPB	20 Jun 2018 10:11:02	U	SMPL
ChkStd - 1	ChkStd	25949	6.4541	-7470A	PPB	20 Jun 2018 10:13:09	U	SMPL
PB110400BL - 1	PBW	-202	-0.0521	-7470A	PPB	20 Jun 2018 10:15:17	U	SMPL
PB110400BS - 1	LCSW	14509	3.6079	-7470A	PPB	20 Jun 2018 10:17:26	U	SMPL
PB110400TB - 1	PB110362TB	-124	-0.0327	-7470A	PPB	20 Jun 2018 10:19:34	U	SMPL
J3498-04 - 1	TP-G-S	-27	-0.0086	-7470A	PPB	20 Jun 2018 10:21:42	U	SMPL
J3498-04DUP - 1	TP-G-SDUP	-19	-0.0066	-7470A	PPB	20 Jun 2018 10:23:50	U	SMPL
J3498-04MS - 1	TP-G-SMS	12660	3.1479	-7470A	PPB	20 Jun 2018 10:25:59	U	SMPL
J3498-04MSD - 1	TP-G-SMSD	13282	3.3026	-7470A	PPB	20 Jun 2018 10:28:06	U	SMPL
CCV82 - 1	CCV82	20646	5.1348	-7470A	PPB	20 Jun 2018 10:30:13	U	SMPL
CCB82 - 1	CCB82	-165	-0.0429	-7470A	PPB	20 Jun 2018 10:32:20	U	SMPL
J3498-08 - 1	TP-G-D	12	0.0011	-7470A	PPB	20 Jun 2018 10:34:27	U	SMPL
J3555-01 - 1	DRUM-1-4-OIL	2299	0.5701	-7470A	PPB	20 Jun 2018 10:36:34	U	SMPL
J3555-02 - 1	DRUM-1-4-OILY-DEBRIS	-5	-0.0031	-7470A	PPB	20 Jun 2018 10:38:41	U	SMPL
J3573-09 - 1	SB-47-COMP	-19	-0.0066	-7470A	PPB	20 Jun 2018 10:40:48	U	SMPL
J3573-10 - 1	SB-48-COMP	-27	-0.0086	-7470A	PPB	20 Jun 2018 10:42:55	U	SMPL
J3573-11 - 1	SB-49-COMP	-6	-0.0034	-7470A	PPB	20 Jun 2018 10:45:02	U	SMPL
J3573-12 - 1	SB-50-COMP	-4	-0.0029	-7470A	PPB	20 Jun 2018 10:47:10	U	SMPL
J3573-13 - 1	SB-54-COMP	2	-0.0014	-7470A	PPB	20 Jun 2018 10:49:18	U	SMPL
J3573-14 - 1	SB-45-COMP	18	0.0026	-7470A	PPB	20 Jun 2018 10:51:26	U	SMPL
J3573-20 - 1	SB-52-COMP	28	0.0051	-7470A	PPB	20 Jun 2018 10:53:33	U	SMPL
CCV83 - 1	CCV83	20383	5.0693	-7470A	PPB	20 Jun 2018 10:55:40	U	SMPL
CCB83 - 1	CCB83	-174	-0.0452	-7470A	PPB	20 Jun 2018 10:57:47	U	SMPL
J3573-21 - 1	SB-53-COMP	11	0.0009	-7470A	PPB	20 Jun 2018 10:59:54	U	SMPL
J3575-03 - 1	TP-9	-17	-0.0061	-7470A	PPB	20 Jun 2018 11:02:01	U	SMPL
J3575-07 - 1	TP-10	35	0.0068	-7470A	PPB	20 Jun 2018 11:04:08	U	SMPL
J3575-11 - 1	TP-12	45	0.0093	-7470A	PPB	20 Jun 2018 11:06:15	U	SMPL
J3575-15 - 1	TP-11-S	86	0.0195	-7470A	PPB	20 Jun 2018 11:08:22	U	SMPL
J3575-19 - 1	TP-11-D	51	0.0108	-7470A	PPB	20 Jun 2018 11:10:29	U	SMPL
J3575-22 - 1	TP-13	20	0.0031	-7470A	PPB	20 Jun 2018 11:12:38	U	SMPL
J3576-01 - 1	N-1366-OILY-DEBRIS	-20	-0.0068	-7470A	PPB	20 Jun 2018 11:14:46	U	SMPL
PB110402BL - 1	PBW	-22	-0.0073	-7470A	PPB	20 Jun 2018 11:16:55	U	SMPL
PB110402BS - 1	LCSW	14379	3.5756	-7470A	PPB	20 Jun 2018 11:19:01	U	SMPL
CCV84 - 1	CCV84	20573	5.1166	-7470A	PPB	20 Jun 2018 11:21:09	U	SMPL
CCB84 - 1	CCB84	-104	-0.0277	-7470A	PPB	20 Jun 2018 11:23:16	U	SMPL
PB110402TB - 1	PB110368TB	-24	-0.0078	-7470A	PPB	20 Jun 2018 11:25:24	U	SMPL
J3249-02 - 1	PL-01-061818-A	-14	-0.0053	-7470A	PPB	20 Jun 2018 11:27:31	U	SMPL
J3249-04 - 1	PL-01-061818-B	-25	-0.0081	-7470A	PPB	20 Jun 2018 11:29:38	U	SMPL
J3249-04DUP - 1	PL-01-061818-BDUP	29	0.0053	-7470A	PPB	20 Jun 2018 11:31:45	U	SMPL
J3249-04MS - 1	PL-01-061818-BMS	13703	3.4074	-7470A	PPB	20 Jun 2018 11:33:53	U	SMPL
J3249-04MSD - 1	PL-01-061818-BMSD	13737	3.4158	-7470A	PPB	20 Jun 2018 11:36:01	U	SMPL
J3249-06 - 1	PL-01-061818-C	-59	-0.0165	-7470A	PPB	20 Jun 2018 11:38:09	U	SMPL
J3249-08 - 1	PL-01-061818-D	13	0.0014	-7470A	PPB	20 Jun 2018 11:40:17	U	SMPL
J3249-10 - 1	PL-01-061818-E	47	0.0098	-7470A	PPB	20 Jun 2018 11:42:25	U	SMPL
PB110403BL - 1	PBW	39	0.0078	-7470A	PPB	20 Jun 2018 11:44:32	U	SMPL
CCV85 - 1	CCV85	20642	5.1338	-7470A	PPB	20 Jun 2018 11:46:40	U	SMPL
CCB85 - 1	CCB85	-88	-0.0238	-7470A	PPB	20 Jun 2018 11:48:46	U	SMPL
PB110403BS - 1	LCSW	13091	3.2551	-7470A	PPB	20 Jun 2018 11:50:55	U	SMPL
PB110403TB - 1	PB110369TB	-57	-0.0160	-7470A	PPB	20 Jun 2018 11:53:01	U	SMPL
J3532-02 - 1	OILY-WATER-DRUMS-1-2	74	0.0165	-7470A	PPB	20 Jun 2018 11:55:09	U	SMPL
J3532-02DUP - 1	OILY-WATER-DRUMS-1-2DUP	82	0.0185	-7470A	PPB	20 Jun 2018 11:57:16	U	SMPL
J3532-02MS - 1	OILY-WATER-DRUMS-1-2MS	13492	3.3549	-7470A	PPB	20 Jun 2018 11:59:23	U	SMPL
J3532-02MSD - 1	OILY-WATER-DRUMS-1-2MSD	13024	3.2384	-7470A	PPB	20 Jun 2018 12:01:31	U	SMPL
PB110426BL - 1	PBW	-187	-0.0484	-7470A	PPB	20 Jun 2018 12:03:40	U	SMPL
PB110426BS - 1	LCSW	13267	3.2989	-7470A	PPB	20 Jun 2018 12:05:49	U	SMPL
J3528-02 - 1	TPTW-01	-27	-0.0086	-7470A	PPB	20 Jun 2018 12:07:58	U	SMPL
J3528-04 - 1	GW-FIELD-BLANK	-70	-0.0193	-7470A	PPB	20 Jun 2018 12:10:05	U	SMPL
CCV86 - 1	CCV86	20609	5.1256	-7470A	PPB	20 Jun 2018 12:12:13	U	SMPL
CCB86 - 1	CCB86	-98	-0.0262	-7470A	PPB	20 Jun 2018 12:14:20	U	SMPL
J3570-05 - 1	20180375-COMPOSITE-CS-1A-ELUTRIATE	123	0.0287	-7470A	PPB	20 Jun 2018 12:16:28	U	SMPL
J3570-05DUP - 1	20180375-COMPOSITE-CS-1A-ELUTRIATEDUP	126	0.0295	-7470A	PPB	20 Jun 2018 12:18:36	U	SMPL
J3570-05MS - 1	20180375-COMPOSITE-CS-1A-ELUTRIATEMS	13821	3.4367	-7470A	PPB	20 Jun 2018 12:20:43	U	SMPL
J3570-05MSD - 1	20180375-COMPOSITE-CS-1A-ELUTRIATEMSD	13599	3.3815	-7470A	PPB	20 Jun 2018 12:22:50	U	SMPL
J3570-06 - 1	20180376-COMPOSITE-CS-1B-ELUTRIATE	9	0.0004	-7470A	PPB	20 Jun 2018 12:24:57	U	SMPL
J3570-12 - 1	20180375-COMPOSITE-CS-1A-ELUTRIATE	-22	-0.0073	-7470A	PPB	20 Jun 2018 12:27:05	U	SMPL
J3570-13 - 1	20180376-COMPOSITE-CS-1B-ELUTRIATE	-38	-0.0113	-7470A	PPB	20 Jun 2018 12:29:13	U	SMPL
CCV87 - 1	CCV87	20800	5.1731	-7470A	PPB	20 Jun 2018 12:31:21	U	SMPL
CCB87 - 1	CCB87	-102	-0.0272	-7470A	PPB	20 Jun 2018 12:33:30	U	SMPL
J3498-04LX5 - 1		-62	-0.0173	-7470A	PPB	20 Jun 2018 12:35:38	U	SMPL
J3498-04A - 1		13904	3.4574	-7470A	PPB	20 Jun 2018 12:37:46	U	SMPL
J3249-04LX5 - 1		-134	-0.0352	-7470A	PPB	20 Jun 2018 12:39:54	U	SMPL
J3249-04A - 1		14205	3.5323	-7470A	PPB	20 Jun 2018 12:42:02	U	SMPL
J3532-02LX5 - 1		-131	-0.0345	-7470A	PPB	20 Jun 2018 12:44:09	U	SMPL
J3532-02A - 1		13327	3.3138	-7470A	PPB	20 Jun 2018 12:46:17	U	SMPL
J3570-05LX5 - 1		-149	-0.0389	-7470A	PPB	20 Jun 2018 12:48:24	U	SMPL
J3570-05A - 1		15926	3.9605	-7470A	PPB	20 Jun 2018 12:50:32	U	SMPL
CCV88 - 1	CCV88	20365	5.0649	-7470A	PPB	20 Jun 2018 12:52:40	U	SMPL
CCB88 - 1	CCB88	-160	-0.0417	-7470A	PPB	20 Jun 2018 12:54:48	U	SMPL