



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB96139

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.3	4.01	06/20/2018	09:30
2	CAL2	1	Water			21.4	7.00	06/20/2018	09:32
3	CAL3	1	Water			21.4	10.03	06/20/2018	09:34
4	ICV	1	Water			21.4	7.00	06/20/2018	09:36
5	CCV1	1	Water			21.3	2.00	06/20/2018	09:38
6	J3603-02	1	Solid	20.02	20	22.3	6.58	06/20/2018	09:40
7	J3603-02DUP	1	Solid	20.02	20	22.4	6.59	06/20/2018	09:42
8	J3603-04	1	Solid	20.04	20	22.2	7.13	06/20/2018	09:44
9	J3603-06	1	Solid	20.03	20	21.6	10.39	06/20/2018	09:46
10	J3603-08	1	Solid	20.02	20	21.6	8.84	06/20/2018	09:48
11	J3603-10	1	Solid	20.05	20	20.9	9.65	06/20/2018	09:50
12	J3603-12	1	Solid	20.02	20	21.6	8.88	06/20/2018	09:52
13	J3603-14	1	Solid	20.03	20	21.2	8.82	06/20/2018	09:54
14	J3603-16	1	Solid	20.03	20	21.2	7.25	06/20/2018	09:56
15	J3603-18	1	Solid	20.02	20	21.1	7.24	06/20/2018	09:58
16	CCV2	1	Water			21.4	12.00	06/20/2018	10:00
17	J3603-20	1	Solid	20.03	20	20.6	6.61	06/20/2018	10:02
18	J3603-22	1	Solid	20.02	20	20.5	7.36	06/20/2018	10:04
19	J3603-22DUP	1	Solid	20.03	20	20.6	7.37	06/20/2018	10:06
20	J3603-24	1	Solid	20.04	20	20.2	6.77	06/20/2018	10:08

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.2	2.00	06/20/2018	10:10

WORKLIST(Hardcopy Internal Chain)

LB96139

WorkList Name : corrosivity062018

WorkList ID : 113582

Date : 6/20/2018 9:14:26 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/19/2018	Solid	J3603-02	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-1	06/19/2018	9045C
06/19/2018	Solid	J3603-04	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-2	06/19/2018	9045C
06/19/2018	Solid	J3603-06	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-3	06/19/2018	9045C
06/19/2018	Solid	J3603-08	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-4	06/19/2018	9045C
06/19/2018	Solid	J3603-10	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-5	06/19/2018	9045C
06/19/2018	Solid	J3603-12	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-CSM	06/19/2018	9045C
06/19/2018	Solid	J3603-14	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-10	06/19/2018	9045C
06/19/2018	Solid	J3603-16	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-9	06/19/2018	9045C
06/19/2018	Solid	J3603-18	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-8	06/19/2018	9045C
06/19/2018	Solid	J3603-20	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-7	06/19/2018	9045C
06/19/2018	Solid	J3603-22	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-6	06/19/2018	9045C
06/19/2018	Solid	J3603-24	Corrosivity	Cool 4 deg C	PSEG01	D62	26KV-TP-11	06/19/2018	9045C

Date/Time 06/20/18 9:20 AM

Received by: SP

Relinquished by: CP

Date/Time 06/20/18 10:15 AM

Received by: CP

Relinquished by: SP