


7471B

LB 96158
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41433



A= 0.0000e+000

B= 2.4486e-004

C= -3.6710e-002

Rho= 0.9996054

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	80	0.000	80				
0.2	0.200	0.200	-0.000	966	0.0 %	966				
2.5	2.500	2.565	0.065	10627	0.0 %	10627				
5.0	5.000	5.055	0.055	20796	0.0 %	20796				
7.5	7.500	7.288	-0.212	29913	0.0 %	29913				
10.0	10.000	10.109	0.109	41433	0.0 %	41433				

LB96158 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Jun 2018 15:37:36

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	80	-	0.00007471B	PPB		20 Jun 2018 15:43:48	S	Std
0.2 - 1	50.5	966	-	0.20007471B	PPB		20 Jun 2018 15:46:05	S	Std
2.5 - 1	52.5	10627	-	2.50007471B	PPB		20 Jun 2018 15:48:23	S	Std
5.0 - 1	55	20796	-	5.00007471B	PPB		20 Jun 2018 15:50:40	S	Std
7.5 - 1	57.5	29913	-	7.50007471B	PPB		20 Jun 2018 15:52:57	S	Std
10.0 - 1	510	41433	-	10.00007471B	PPB		20 Jun 2018 15:58:02	S	Std
ICV12 - 1	ICV12	16353	3.9675	-7471B	PPB		20 Jun 2018 16:01:00	U	SMPL
ICB12 - 1	ICB12	-47	-0.0482	-7471B	PPB		20 Jun 2018 16:03:16	U	SMPL
CCV89 - 1	CCV89	21115	5.1335	-7471B	PPB		20 Jun 2018 16:05:32	U	SMPL
CCB89 - 1	CCB89	-67	-0.0531	-7471B	PPB		20 Jun 2018 16:07:48	U	SMPL
CRA - 1	CRA	968	0.2003	-7471B	PPB		20 Jun 2018 16:10:05	U	SMPL
HighStd - 1	HighStd	42397	10.3447	-7471B	PPB		20 Jun 2018 16:12:21	U	SMPL
ChkStd - 1	ChkStd	28404	6.9183	-7471B	PPB		20 Jun 2018 16:14:37	U	SMPL
PB110412BL - 1	PBS	-114	-0.0646	-7471B	PPB		20 Jun 2018 16:16:55	U	SMPL
PB110412BS - 1	LCSS	16084	3.9016	-7471B	PPB		20 Jun 2018 16:19:13	U	SMPL
J3245-09 - 1	CL-02-061918-A	2179	0.4968	-7471B	PPB		20 Jun 2018 16:21:30	U	SMPL
J3245-09DUP - 1	CL-02-061918-ADUP	1816	0.4080	-7471B	PPB		20 Jun 2018 16:30:54	U	SMPL
J3245-09MS - 1	CL-02-061918-AMS	18562	4.5084	-7471B	PPB		20 Jun 2018 16:33:11	U	SMPL
J3245-09MSD - 1	CL-02-061918-AMSD	18509	4.4954	-7471B	PPB		20 Jun 2018 16:35:29	U	SMPL
J3245-11 - 1	CL-02-061918-B	1127	0.2392	-7471B	PPB		20 Jun 2018 16:37:44	U	SMPL
CCV90 - 1	CCV90	20648	5.0192	-7471B	PPB		20 Jun 2018 16:40:01	U	SMPL
CCB90 - 1	CCB90	-63	-0.0521	-7471B	PPB		20 Jun 2018 16:42:16	U	SMPL
J3245-13 - 1	CL-02-061918-C	2285	0.5228	-7471B	PPB		20 Jun 2018 16:44:33	U	SMPL
J3245-15 - 1	CL-02-061918-D	2608	0.6019	-7471B	PPB		20 Jun 2018 16:46:49	U	SMPL
J3247-05 - 1	OR-01-061918	1834	0.4124	-7471B	PPB		20 Jun 2018 16:49:05	U	SMPL
J3589-01 - 1	B-1	234	0.0206	-7471B	PPB		20 Jun 2018 16:51:22	U	SMPL
J3589-02 - 1	B-2	1288	0.2787	-7471B	PPB		20 Jun 2018 16:53:38	U	SMPL
J3589-03 - 1	B-3	1271	0.2745	-7471B	PPB		20 Jun 2018 16:55:56	U	SMPL
J3593-01 - 1	S39-9005	241	0.0223	-7471B	PPB		20 Jun 2018 16:58:13	U	SMPL
J3594-01 - 1	HC-SOIL-1	1658	0.3693	-7471B	PPB		20 Jun 2018 17:00:31	U	SMPL
J3594-03 - 1	HC-SOIL-2	1677	0.3739	-7471B	PPB		20 Jun 2018 17:02:49	U	SMPL
J3594-05 - 1	HC-SOIL-3	1468	0.3227	-7471B	PPB		20 Jun 2018 17:05:05	U	SMPL
CCV91 - 1	CCV91	20613	5.0106	-7471B	PPB		20 Jun 2018 17:07:21	U	SMPL
CCB91 - 1	CCB91	-103	-0.0619	-7471B	PPB		20 Jun 2018 17:09:37	U	SMPL
J3594-07 - 1	HC-SOIL-4	1317	0.2858	-7471B	PPB		20 Jun 2018 17:11:54	U	SMPL
J3595-01 - 1	SC-01-ABC	2361	0.5414	-7471B	PPB		20 Jun 2018 17:14:10	U	SMPL
J3595-03 - 1	SC-02-AB	33670	8.2078	-7471B	PPB		20 Jun 2018 17:16:26	U	SMPL
J3598-01 - 1	S73-9001	0	-0.0367	-7471B	PPB		20 Jun 2018 17:18:43	U	SMPL
J3600-01 - 1	DOCUMENTATION-1	33182	8.0883	-7471B	PPB		20 Jun 2018 17:21:00	U	SMPL
J3600-02 - 1	DOCUMENTATION-2	22926	5.5770	-7471B	PPB		20 Jun 2018 17:23:17	U	SMPL
J3600-03 - 1	DOCUMENTATION-3	20024	4.8664	-7471B	PPB		20 Jun 2018 17:25:36	U	SMPL
PB110413BL - 1	PBS	-84	-0.0573	-7471B	PPB		20 Jun 2018 17:27:54	U	SMPL
PB110413BS - 1	LCSS	16508	4.0055	-7471B	PPB		20 Jun 2018 17:30:11	U	SMPL
J3602-01 - 1	EPE-F-8(0-5)	2061	0.4679	-7471B	PPB		20 Jun 2018 17:32:28	U	SMPL
CCV92 - 1	CCV92	20620	5.0123	-7471B	PPB		20 Jun 2018 17:34:44	U	SMPL
CCB92 - 1	CCB92	-112	-0.0641	-7471B	PPB		20 Jun 2018 17:37:01	U	SMPL
J3602-02 - 1	EPE-F-8(5-10)	686	0.1313	-7471B	PPB		20 Jun 2018 17:39:18	U	SMPL
J3603-01 - 1	26KV-TP-1	20417	4.9626	-7471B	PPB		20 Jun 2018 17:41:34	U	SMPL
J3603-03 - 1	26KV-TP-2	107968	26.4005	-7471B	PPB		20 Jun 2018 17:43:50	U	SMPL
J3603-05 - 1	26KV-TP-3	15658	3.7973	-7471B	PPB		20 Jun 2018 17:46:07	U	SMPL
J3603-07 - 1	26KV-TP-4	35582	8.6759	-7471B	PPB		20 Jun 2018 17:48:26	U	SMPL
J3603-09 - 1	26KV-TP-5	20830	5.0637	-7471B	PPB		20 Jun 2018 17:50:42	U	SMPL
J3603-11 - 1	26KV-CSM	4	-0.0357	-7471B	PPB		20 Jun 2018 17:53:01	U	SMPL
J3603-13 - 1	26KV-TP-10	18926	4.5975	-7471B	PPB		20 Jun 2018 17:55:19	U	SMPL
J3603-15 - 1	26KV-TP-9	38822	9.4693	-7471B	PPB		20 Jun 2018 17:57:37	U	SMPL
J3603-17 - 1	26KV-TP-8	28932	7.0476	-7471B	PPB		20 Jun 2018 17:59:54	U	SMPL
CCV93 - 1	CCV93	20045	4.8715	-7471B	PPB		20 Jun 2018 18:02:11	U	SMPL
CCB93 - 1	CCB93	-118	-0.0656	-7471B	PPB		20 Jun 2018 18:04:29	U	SMPL
J3603-19 - 1	26KV-TP-7	12814	3.1009	-7471B	PPB		20 Jun 2018 18:06:46	U	SMPL
J3603-21 - 1	26KV-TP-6	46533	11.3574	-7471B	PPB		20 Jun 2018 18:09:02	U	SMPL
J3603-23 - 1	26KV-TP-11	35762	8.7200	-7471B	PPB		20 Jun 2018 18:11:19	U	SMPL
J3606-01 - 1	FK-PS-01-030	310	0.0392	-7471B	PPB		20 Jun 2018 18:13:37	U	SMPL
J3606-01DUP - 1	FK-PS-01-030DUP	307	0.0385	-7471B	PPB		20 Jun 2018 18:21:16	U	SMPL
J3606-01MS - 1	FK-PS-01-030MS	16977	4.1203	-7471B	PPB		20 Jun 2018 18:23:32	U	SMPL
J3606-01MSD - 1	FK-PS-01-030MSD	17079	4.1453	-7471B	PPB		20 Jun 2018 18:25:50	U	SMPL
CCV94 - 1	CCV94	20247	4.9210	-7471B	PPB		20 Jun 2018 18:28:07	U	SMPL
CCB94 - 1	CCB94	-113	-0.0644	-7471B	PPB		20 Jun 2018 18:30:25	U	SMPL
J3245-09LX5 - 1		343	0.0473	-7471B	PPB		20 Jun 2018 18:32:42	U	SMPL
J3245-09A - 1		18601	4.5180	-7471B	PPB		20 Jun 2018 18:34:58	U	SMPL
J3606-01LX5 - 1		-55	-0.0502	-7471B	PPB		20 Jun 2018 18:37:15	U	SMPL
J3606-01A - 1		17144	4.1612	-7471B	PPB		20 Jun 2018 18:39:32	U	SMPL
J3603-03DLX5 - 1		22029	5.3573	-7471B	PPB		20 Jun 2018 18:41:48	U	SMPL
J3603-21DLX2 - 1		24138	5.8737	-7471B	PPB		20 Jun 2018 19:13:34	U	SMPL
CCV95 - 1	CCV95	20714	5.0353	-7471B	PPB		20 Jun 2018 19:15:50	U	SMPL
CCB95 - 1	CCB95	-128	-0.0681	-7471B	PPB		20 Jun 2018 19:18:07	U	SMPL