



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB96162

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2220
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.4	4.01	06/20/2018	15:30
2	CAL2	1	Water			21.3	7.00	06/20/2018	15:32
3	CAL3	1	Water			21.5	10.03	06/20/2018	15:34
4	ICV	1	Water			21.3	7.00	06/20/2018	15:36
5	CCV1	1	Water			21.4	2.00	06/20/2018	15:38
6	J3628-01	1	Solid	20.02	20	22.9	6.37	06/20/2018	15:40
7	J3628-01DUP	1	Solid	20.03	20	23.0	6.38	06/20/2018	15:42
8	J3629-01	1	Solid	20.02	20	22.4	4.73	06/20/2018	15:44
9	J3629-02	1	Solid	20.03	20	22.4	6.41	06/20/2018	15:46
10	J3629-05	1	Solid	20.02	20	22.2	8.80	06/20/2018	15:48
11	J3629-06	1	Solid	20.03	20	21.9	8.97	06/20/2018	15:50
12	J3629-07	1	Solid	20.02	20	22.0	9.21	06/20/2018	15:52
13	J3629-08	1	Solid	20.04	20	21.7	9.53	06/20/2018	15:54
14	J3629-09	1	Solid	20.02	20	21.7	9.50	06/20/2018	15:56
15	J3629-13	1	Solid	20.04	20	21.3	8.19	06/20/2018	15:58
16	CCV2	1	Water			21.4	12.00	06/20/2018	16:00
17	J3629-14	1	Solid	20.02	20	21.6	9.01	06/20/2018	16:02
18	J3629-15	1	Solid	20.03	20	21.3	7.67	06/20/2018	16:04
19	J3629-16	1	Solid	20.02	20	21.1	7.01	06/20/2018	16:06
20	J3629-16DUP	1	Solid	20.04	20	21.2	7.02	06/20/2018	16:08

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water			21.3	2.00	06/20/2018	16:10

WORKLIST(Hardcopy Internal Chain)

LB96162

WorkList Name : pH062018

WorkList ID : 113604

Date : 6/20/2018 3:10:53 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
06/18/2018	Solid	J3628-01	pH	Cool 4 deg C	AREC01	D62	A0C11-01	06/18/2018	9045C
06/18/2018	Solid	J3629-01	pH	Cool 4 deg C	AREC01	D31	A0C2-01	06/18/2018	9045C
06/18/2018	Solid	J3629-02	pH	Cool 4 deg C	AREC01	D31	A0C2-02	06/18/2018	9045C
06/18/2018	Solid	J3629-05	pH	Cool 4 deg C	AREC01	D31	A0C3-01	06/18/2018	9045C
06/18/2018	Solid	J3629-06	pH	Cool 4 deg C	AREC01	D31	A0C10-01	06/18/2018	9045C
06/18/2018	Solid	J3629-07	pH	Cool 4 deg C	AREC01	D31	A0C10-02	06/18/2018	9045C
06/18/2018	Solid	J3629-08	pH	Cool 4 deg C	AREC01	D31	A0C10-03	06/18/2018	9045C
06/18/2018	Solid	J3629-09	pH	Cool 4 deg C	AREC01	D31	A0C10-04	06/18/2018	9045C
06/18/2018	Solid	J3629-13	pH	Cool 4 deg C	AREC01	D31	A0C14-01	06/18/2018	9045C
06/18/2018	Solid	J3629-14	pH	Cool 4 deg C	AREC01	D31	A0C14-02	06/18/2018	9045C
06/18/2018	Solid	J3629-15	pH	Cool 4 deg C	AREC01	D31	A0C22-01	06/18/2018	9045C
06/18/2018	Solid	J3629-16	pH	Cool 4 deg C	AREC01	D31	A0C22-02	06/18/2018	9045C

Date/Time 06/20/18 3:15 PM

Received by: SP

Relinquished by: CP

Date/Time 06/20/18 4:15 PM

Received by: CP

Relinquished by: SP